


LB109248

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	159	0.000	159				
0.2	0.200	0.202	0.002	926	0.0 %	926				
2.5	2.500	2.535	0.035	9916	0.0 %	9916				
5.0	5.000	4.982	-0.018	19348	0.0 %	19348				
7.5	7.500	7.414	-0.086	28722	0.0 %	28722				
10.0	10.000	10.064	0.064	38937	0.0 %	38937				

LB109248 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 May 2020 15:31:44

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.0000	7470A	PPB	20 May 2020 15:37:46	S	Std
0.2 - 1	50.2	926	-	0.2000	7470A	PPB	20 May 2020 15:39:53	S	Std
2.5 - 1	52.5	9916	-	2.5000	7470A	PPB	20 May 2020 15:42:01	S	Std
5.0 - 1	55	19348	-	5.0000	7470A	PPB	20 May 2020 15:44:09	S	Std
7.5 - 1	57.5	28722	-	7.5000	7470A	PPB	20 May 2020 15:46:16	S	Std
10.0 - 1	510	38937	-	10.0000	7470A	PPB	20 May 2020 15:48:23	S	Std
ICV55 - 1	ICV55	14553	3.7378	-	7470A	PPB	20 May 2020 15:52:35	U	SMPL
ICB55 - 1	ICB55	40	-0.0274	-	7470A	PPB	20 May 2020 15:54:40	U	SMPL
CCV78 - 1	CCV78	19087	4.9140	-	7470A	PPB	20 May 2020 15:56:46	U	SMPL
CCB78 - 1	CCB78	34	-0.0289	-	7470A	PPB	20 May 2020 15:58:51	U	SMPL
CRA - 1	CRA	917	0.2001	-	7470A	PPB	20 May 2020 16:00:57	U	SMPL
HighStd - 1	HighStd	38778	10.0226	-	7470A	PPB	20 May 2020 16:03:03	U	SMPL
ChkStd - 1	ChkStd	25513	6.5812	-	7470A	PPB	20 May 2020 16:05:09	U	SMPL
PB129017BL - 1	PBW	-53	-0.0515	-	7470A	PPB	20 May 2020 16:07:17	U	SMPL
PB129017BS - 1	LCSW	15197	3.9049	-	7470A	PPB	20 May 2020 16:09:24	U	SMPL
L2497-08 - 1	HD-02-051920	23	-0.0318	-	7470A	PPB	20 May 2020 16:11:31	U	SMPL
L2497-08DUP - 1	HD-02-051920DUP	141	-0.0012	-	7470A	PPB	20 May 2020 16:13:39	U	SMPL
L2497-08MS - 1	HD-02-051920MS	15175	3.8991	-	7470A	PPB	20 May 2020 16:15:46	U	SMPL
L2497-08MSD - 1	HD-02-051920MSD	15291	3.9292	-	7470A	PPB	20 May 2020 16:17:54	U	SMPL
L2713-01 - 1	BUS-WASH-SLUDGE-PIT	85	-0.0157	-	7470A	PPB	20 May 2020 16:20:00	U	SMPL
CCV79 - 1	CCV79	19188	4.9403	-	7470A	PPB	20 May 2020 16:22:06	U	SMPL
CCB79 - 1	CCB79	37	-0.0282	-	7470A	PPB	20 May 2020 16:24:12	U	SMPL
L2713-02 - 1	ZIPPER-DRAINS-SLUDGE-PIT	299	0.0398	-	7470A	PPB	20 May 2020 16:26:18	U	SMPL
PB129018BL - 1	PBW	102	-0.0113	-	7470A	PPB	20 May 2020 16:28:25	U	SMPL
PB129018BS - 1	LCSW	15835	4.0704	-	7470A	PPB	20 May 2020 16:30:31	U	SMPL
L2714-01 - 1	RP-WATER	84	-0.0160	-	7470A	PPB	20 May 2020 16:32:37	U	SMPL
L2714-01DUP - 1	RP-WATERDUP	149	0.0009	-	7470A	PPB	20 May 2020 16:34:44	U	SMPL
L2714-01MS - 1	RP-WATERMS	17063	4.3890	-	7470A	PPB	20 May 2020 16:36:51	U	SMPL
L2714-01MSD - 1	RP-WATERMSD	17190	4.4219	-	7470A	PPB	20 May 2020 16:38:58	U	SMPL
CCV80 - 1	CCV80	19221	4.9488	-	7470A	PPB	20 May 2020 16:41:06	U	SMPL
CCB80 - 1	CCB80	38	-0.0279	-	7470A	PPB	20 May 2020 16:43:14	U	SMPL
PB129017TB - 1	PB129001TB	110	-0.0092	-	7470A	PPB	20 May 2020 16:45:21	U	SMPL
L2497-08LX5 - 1		126	-0.0051	-	7470A	PPB	20 May 2020 16:47:27	U	SMPL
L2497-08A - 1		15174	3.8989	-	7470A	PPB	20 May 2020 16:49:33	U	SMPL
L2714-01LX5 - 1		37	-0.0282	-	7470A	PPB	20 May 2020 16:51:40	U	SMPL
L2714-01A - 1		17060	4.3882	-	7470A	PPB	20 May 2020 16:53:46	U	SMPL
CCV81 - 1	CCV81	19415	4.9991	-	7470A	PPB	20 May 2020 16:55:52	U	SMPL
CCB81 - 1	CCB81	21	-0.0323	-	7470A	PPB	20 May 2020 16:57:59	U	SMPL