



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109269

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	05/21/2020	16:00
2	CAL2	1	Water	NA	NA	20.2	7.01	05/21/2020	16:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/21/2020	16:03
4	ICV	1	Water	NA	NA	20.1	7.00	05/21/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/21/2020	16:10
6	L2738-01	1	Water	NA	NA	16.3	6.38	05/21/2020	16:12
7	L2738-01DUP	1	Water	NA	NA	16.4	6.39	05/21/2020	16:13
8	L2738-02	1	Water	NA	NA	16.0	6.14	05/21/2020	16:15
9	CCV2	1	Water	NA	NA	20.2	12.02	05/21/2020	16:16

LB 109269

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2738 WorkList ID : 138963 Department : Wet-Chemistry Date : 05-21-2020 15:42:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2020	Water	L2738-01 ✓	pH	Cool 4 deg C	PSEG01	N21	1725	05/21/2020	9040C
05/21/2020	Water	L2738-02 ✓	pH	Cool 4 deg C	PSEG01	N21	1728	05/21/2020	9040C

Date/Time 05-21-2020 15:50
 Received by: JS
 Relinquished by:

Date/Time 05-21-2020 17:40
 Received by: JS
 Relinquished by: JS