



Analytical Summary Report

Analysis Method: 9041A

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109270

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616
3.0-5.5 pH paper refill	W1934
Test Paper, PH, Short Range 6.0/8.0	W2350
10.0-12.0 pH paper refill	W1940
0.0-3.0 pH paper refill	W1931
10.0-12.0 pH paper refill	W1940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.0	05/21/2020	16:25
2	CAL2	1	Water	NA	NA	20.1	7.0	05/21/2020	16:26
3	CAL3	1	Water	NA	NA	20.1	10.0	05/21/2020	16:27
4	ICV	1	Water	NA	NA	20.2	7.0	05/21/2020	16:30
5	CCV1	1	Water	NA	NA	20.1	2.0	05/21/2020	16:31
6	L2738-12	1	Solid	20.02	20	21.1	5.5	05/21/2020	16:33
7	L2738-15	1	Solid	20.03	20	21.3	5.6	05/21/2020	16:35
8	L2738-15DUP	1	Solid	20.02	20	21.4	5.6	05/21/2020	16:36
9	CCV2	1	Water	NA	NA	20.1	12.0	05/21/2020	16:37

13109270

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH S L2738 WorkList ID : 138964 Department : Wet-Chemistry Date : 05-21-2020 15:43:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2020	Solid	L2738-12 - C	pH	Cool 4 deg C	PSEG01	N21	1702	05/21/2020	9041A
05/21/2020	Solid	L2738-15 - C	pH	Cool 4 deg C	PSEG01	N21	1723	05/21/2020	9041A

Date/Time 05.21.2020 15150
 Received by: KS
 Relinquished by: KS

Date/Time 05.21.2020 17100
 Received by: KS
 Relinquished by: KS