



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109287

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2628
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/22/2020	15:00
2	CAL2	1	Water	NA	NA	20.2	7.02	05/22/2020	15:01
3	CAL3	1	Water	NA	NA	20.2	10.03	05/22/2020	15:03
4	ICV	1	Water	NA	NA	20.1	7.00	05/22/2020	15:05
5	CCV1	1	Water	NA	NA	20.3	2.01	05/22/2020	15:06
6	L2752-01	1	Water	NA	NA	21.6	5.38	05/22/2020	15:10
7	L2752-02	1	Water	NA	NA	22.0	5.42	05/22/2020	15:12
8	L2752-03	1	Water	NA	NA	22.4	5.40	05/22/2020	15:15
9	L2752-03DUP	1	Water	NA	NA	22.5	5.41	05/22/2020	15:16
10	CCV2	1	Water	NA	NA	20.1	12.02	05/22/2020	15:19

VB109287

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph I2732 WorkList ID : 139011 Department : Wet-Chemistry Date : 05-22-2020 14:46:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/21/2020	Water	L2752-01	pH -A	Cool 4 deg C	STER13	N22	CELL-A-POOL-WATER	05/21/2020	SM 4500-H B
05/21/2020	Water	L2752-02	pH -A	Cool 4 deg C	STER13	N22	CELL-B-POOL-WATER	05/21/2020	SM 4500-H B
05/21/2020	Water	L2752-03	pH -A	Cool 4 deg C	STER13	N22	CELL-C-POOL-WATER	05/21/2020	SM 4500-H B

Date/Time 05.22.2020 14:50
Received by: Jo
Relinquished by: Du

Date/Time 05.22.2020 16:10
Received by: Du
Relinquished by: JD