



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109291

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/22/2020	09:20
2	L2738-12		1	8	100.00	>212.0	>212.0	05/22/2020	09:50
3	L2738-12DUP		1	8	100.00	>212.0	>212.0	05/22/2020	10:25
4	L2738-15		1	9	100.00	>212.0	>212.0	05/22/2020	10:55

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp52120 WorkList ID : 138947 Department : Wet-Chemistry Date : 05-21-2020 09:15:09

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2020	Solid	L2738-12	Flash Point	Cool 4 deg C	PSEG01	N21	1702	05/21/2020	1010A
05/28/2020	Solid	L2738-15	Flash Point	Cool 4 deg C	PSEG01	N21	1723	05/21/2020	1010A

Date/Time 05/22/20 8:45
 Received by: KS
 Relinquished by: JP

Date/Time 05/22/20 11:00
 Received by: JP
 Relinquished by: KS