



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109324

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	05/27/2020	09:25
2	CAL2	1	Water	NA	NA	21.3	7.00	05/27/2020	09:26
3	CAL3	1	Water	NA	NA	21.4	10.02	05/27/2020	09:27
4	ICV	1	Water	NA	NA	21.3	6.99	05/27/2020	09:30
5	CCV1	1	Water	NA	NA	20.1	2.02	05/27/2020	09:32
6	L2767-03	1	Water	NA	NA	23.8	7.02	05/27/2020	09:33
7	L2767-03DUP	1	Water	NA	NA	23.9	7.03	05/27/2020	09:34
8	CCV2	1	Water	NA	NA	20.1	12.02	05/27/2020	09:35

VR 109324

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph w 12767 WorkList ID : 139076 Department : Wet-Chemistry Date : 05-27-2020 09:18:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/26/2020	Water	L2767-03	F pH	Cool 4 deg C	PSEG01	N23	TRE-1006	05/26/2020	9041A

Date/Time 05/27/2020 9:20
Received by: JS
Relinquished by: JS

Date/Time 05/27/2020 10:30
Received by: KS
Relinquished by: JS