



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109328

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/27/2020	11:05
2	CAL2	1	Water	NA	NA	21.2	7.00	05/27/2020	11:06
3	CAL3	1	Water	NA	NA	21.2	10.03	05/27/2020	11:07
4	ICV	1	Water	NA	NA	21.3	7.02	05/27/2020	11:10
5	CCV1	1	Water	NA	NA	20.1	2.02	05/27/2020	11:11
6	L2772-01	1	Water	NA	NA	13.6	6.64	05/27/2020	11:12
7	L2772-01DUP	1	Water	NA	NA	13.7	6.65	05/27/2020	11:12
8	CCV2	1	Water	NA	NA	20.1	12.02	05/27/2020	11:15

WORKLIST(Hardcopy Internal Chain)

LB 109328

WorkList Name : PH L2772 WorkList ID : 139088 Department : Wet-Chemistry Date : 05-27-2020 10:35:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/26/2020	Water	L2772-01	pH	Cool 4 deg C	APT101	N11	WGGM-20-WWW-TSF-018	05/26/2020	SM 4500-H B

Date/Time 05/26/2020 11:00
 Received by: JP
 Relinquished by: JP

Date/Time 05/27/2020 12:00
 Received by: JP
 Relinquished by: JP