



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109347

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	05/28/2020	14:00
2	CAL2	1	Water	NA	NA	21.5	7.01	05/28/2020	14:01
3	CAL3	1	Water	NA	NA	21.6	10.03	05/28/2020	14:03
4	ICV	1	Water	NA	NA	21.6	7.02	05/28/2020	14:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/28/2020	14:06
6	L2795-07	1	Water	NA	NA	24.9	7.27	05/28/2020	14:10
7	L2795-07DUP	1	Water	NA	NA	24.8	7.28	05/28/2020	14:11
8	L2795-08	1	Water	NA	NA	24.9	7.38	05/28/2020	14:12
9	CCV2	1	Water	NA	NA	20.2	12.02	05/28/2020	14:15

LB 109347

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2795 WorkList ID : 139129 Department : Wet-Chemistry Date : 05-28-2020 13:36:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2020	Water	L2795-07	pH	Cool 4 deg C	PSEG01	N33	214	05/28/2020	9040C
05/28/2020	Water	L2795-08	pH	Cool 4 deg C	PSEG01	N33	215	05/28/2020	9040C

Date/Time 05/28/2020 13:40
Received by: Jo
Relinquished by: Jo

Date/Time 05/28/2020 15:00
Received by: Jo
Relinquished by: Jo