



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109349

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/28/2020	09:25
2	L2767-03		1	8	100.00	>212.0	>212.0	05/28/2020	09:50
3	L2767-03DUP		1	8	100.00	>212.0	>212.0	05/28/2020	10:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp52820 WorkList ID : 139115 Department : Wet-Chemistry Date : 05-28-2020 08:40:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2020	Water	L2767-03	E	Cool 4 deg C	PSEG01	N23	TRE-1006	05/26/2020	1010A

Date/Time 05/28/2020 8:50
 Received by: KS
 Relinquished by: Dh

Date/Time 05/28/2020 10:25
 Received by: Dh
 Relinquished by: KS