


LB109360

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.010	-0.010	29	0.000	29				
0.2	0.200	0.208	0.008	1261	0.0 %	1261				
2.5	2.500	2.524	0.024	14311	0.0 %	14311				
5.0	5.000	4.967	-0.033	28075	0.0 %	28075				
7.5	7.500	7.502	0.002	42357	0.0 %	42357				
10.0	10.000	10.008	0.008	56477	0.0 %	56477				

LB109360 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 May 2020 09:56:09					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.0000	7471B	PPB	29 May 2020 10:00:28	S	Std
0.2 - 1	50.2	1261	-	0.2000	7471B	PPB	29 May 2020 10:02:45	S	Std
2.5 - 1	52.5	14311	-	2.5000	7471B	PPB	29 May 2020 10:05:02	S	Std
5.0 - 1	55	28075	-	5.0000	7471B	PPB	29 May 2020 10:07:18	S	Std
7.5 - 1	57.5	42357	-	7.5000	7471B	PPB	29 May 2020 10:09:35	S	Std
10.0 - 1	510	56477	-	10.0000	7471B	PPB	29 May 2020 10:11:51	S	Std
ICV63 - 1	ICV63	21531	3.8059	-	7471B	PPB	29 May 2020 10:14:35	U	SMPL
ICB63 - 1	ICB63	-64	-0.0269	-	7471B	PPB	29 May 2020 10:16:49	U	SMPL
CCV23 - 1	CCV23	28045	4.9620	-	7471B	PPB	29 May 2020 10:19:05	U	SMPL
CCB23 - 1	CCB23	-77	-0.0292	-	7471B	PPB	29 May 2020 10:21:20	U	SMPL
CRA - 1	CRA	1238	0.2042	-	7471B	PPB	29 May 2020 10:23:35	U	SMPL
HighStd - 1	HighStd	55831	9.8936	-	7471B	PPB	29 May 2020 10:25:50	U	SMPL
ChkStd - 1	ChkStd	39134	6.9301	-	7471B	PPB	29 May 2020 10:28:05	U	SMPL
PB129162BL - 1	PBS	-174	-0.0464	-	7471B	PPB	29 May 2020 10:30:22	U	SMPL
PB129162BS - 1	LCSS	20297	3.5868	-	7471B	PPB	29 May 2020 10:32:39	U	SMPL
L2503-11 - 1	OR-03-052820	3076	0.5304	-	7471B	PPB	29 May 2020 10:34:56	U	SMPL
L2790-03 - 1	SB-1-S	4254	0.7395	-	7471B	PPB	29 May 2020 10:37:12	U	SMPL
L2790-07 - 1	SB-1-D	5601	0.9785	-	7471B	PPB	29 May 2020 10:39:28	U	SMPL
L2790-11 - 1	SB-3-S	3166	0.5464	-	7471B	PPB	29 May 2020 10:41:45	U	SMPL
L2790-17 - 1	SB-7-S	2210	0.3767	-	7471B	PPB	29 May 2020 10:44:00	U	SMPL
CCV24 - 1	CCV24	27948	4.9448	-	7471B	PPB	29 May 2020 10:46:16	U	SMPL
CCB24 - 1	CCB24	-94	-0.0322	-	7471B	PPB	29 May 2020 10:48:31	U	SMPL
L2792-01 - 1	NM32-COMP-2020-0-6	14156	2.4969	-	7471B	PPB	29 May 2020 10:50:47	U	SMPL
L2792-02 - 1	NM28-COMP-2020-0-6	13101	2.3097	-	7471B	PPB	29 May 2020 10:53:02	U	SMPL
L2792-02DUP - 1	NM28-COMP-2020-0-6DUP	15223	2.6863	-	7471B	PPB	29 May 2020 10:55:18	U	SMPL
L2792-03 - 1	L2792-02MS	35975	6.3695	-	7471B	PPB	29 May 2020 10:57:33	U	SMPL
L2792-04 - 1	L2792-02MSD	35708	6.3221	-	7471B	PPB	29 May 2020 10:59:48	U	SMPL
L2792-05 - 1	NM15-COMP-2020-0-6	11199	1.9721	-	7471B	PPB	29 May 2020 11:02:06	U	SMPL
L2792-06 - 1	NM14-COMP-2020-0-6	9333	1.6409	-	7471B	PPB	29 May 2020 11:04:23	U	SMPL
L2792-07 - 1	NM107-COMP-2020-0-6	7631	1.3388	-	7471B	PPB	29 May 2020 11:06:40	U	SMPL
L2792-08 - 1	NM114-COMP-2020-0-6	35727	6.3254	-	7471B	PPB	29 May 2020 11:08:57	U	SMPL
L2792-09 - 1	NM84-COMP-2020-0-6	9895	1.7407	-	7471B	PPB	29 May 2020 11:11:13	U	SMPL
CCV25 - 1	CCV25	27520	4.8688	-	7471B	PPB	29 May 2020 11:13:30	U	SMPL
CCB25 - 1	CCB25	-110	-0.0351	-	7471B	PPB	29 May 2020 11:15:45	U	SMPL
L2792-10 - 1	NM96-COMP-2020-0-6	10182	1.7916	-	7471B	PPB	29 May 2020 11:18:02	U	SMPL
L2792-11 - 1	NM85-COMP-2020-0-6	8431	1.4808	-	7471B	PPB	29 May 2020 11:20:17	U	SMPL
L2792-12 - 1	NM75-COMP-2020-0-6	13028	2.2967	-	7471B	PPB	29 May 2020 11:22:33	U	SMPL
L2792-13 - 1	NM97-COMP-2020-0-6	13710	2.4178	-	7471B	PPB	29 May 2020 11:24:48	U	SMPL
L2792-02LX5 - 1		2471	0.4230	-	7471B	PPB	29 May 2020 11:27:04	U	SMPL
L2792-02A - 1		35566	6.2969	-	7471B	PPB	29 May 2020 11:29:21	U	SMPL
CCV26 - 1	CCV26	27524	4.8695	-	7471B	PPB	29 May 2020 11:31:37	U	SMPL
CCB26 - 1	CCB26	-92	-0.0319	-	7471B	PPB	29 May 2020 11:33:56	U	SMPL