



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109361

Slope : 99.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	05/28/2020	17:00
2	CAL2	1	Water	NA	NA	21.5	7.01	05/28/2020	17:01
3	CAL3	1	Water	NA	NA	21.6	10.03	05/28/2020	17:02
4	ICV	1	Water	NA	NA	21.5	7.01	05/28/2020	17:05
5	CCV1	1	Water	NA	NA	21.6	2.01	05/28/2020	17:06
6	L2796-06	1	Water	NA	NA	23.4	7.92	05/28/2020	17:10
7	L2796-06DUP	1	Water	NA	NA	23.5	7.93	05/28/2020	17:11
8	CCV2	1	Water	NA	NA	21.6	12.02	05/28/2020	17:12

VB109361

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2796 WorkList ID : 139131 Department : Wet-Chemistry Date : 05-28-2020 16:30:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/28/2020	Water	L2796-06	pH	Cool 4 deg C	PSEG01	N33	N48963	05/28/2020	9040C

Date/Time 05/28/2020 16:35
Received by: Jn KS
Relinquished by: KS

Date/Time 05/28/2020 18:00
Received by: KS
Relinquished by: Jn