



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109365

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	05/29/2020	09:10
2	L2795-05		1	8	100.00	>212.0	>212.0	05/29/2020	09:40
3	L2795-05DUP		1	8	100.00	>212.0	>212.0	05/29/2020	10:10
4	L2796-07		1	9	100.00	>212.0	>212.0	05/29/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB109365

WorkList Name : fp52920 WorkList ID : 139158 Department : Wet-Chemistry Date : 05-29-2020 08:30:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2795-05	E Flash Point	Cool 4 deg C	PSEG01	N33	WATER-COMP-214-215	05/28/2020	1010A
06/02/2020	Water	L2796-07	C Flash Point	Cool 4 deg C	PSEG01	N33	N48963	05/28/2020	1010A

Date/Time 05/29/20 08:40
 Received by: KS
 Relinquished by: JP

Date/Time 05/29/20 10:50
 Received by: JD
 Relinquished by: KS