


LB109367

7471B

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.039	0.039	159	0.000	159				
0.2	0.200	0.195	-0.005	1033	0.0 %	1033				
2.5	2.500	2.493	-0.007	13883	0.0 %	13883				
5.0	5.000	4.939	-0.061	27558	0.0 %	27558				
7.5	7.500	7.506	0.006	41915	0.0 %	41915				
10.0	10.000	10.027	0.027	56011	0.0 %	56011				

LB109367 INSTRUMENT ID : CV1

Method: 7471B		Operator: Admin		Date of Analysis: 29 May 2020 12:20:37					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	159	-	0.0000	7471B	PPB	29 May 2020 13:37:18	S	Std
0.2 - 1	50.2	1033	-	0.2000	7471B	PPB	29 May 2020 13:39:35	S	Std
2.5 - 1	52.5	13883	-	2.5000	7471B	PPB	29 May 2020 13:41:52	S	Std
5.0 - 1	55	27558	-	5.0000	7471B	PPB	29 May 2020 13:44:09	S	Std
7.5 - 1	57.5	41915	-	7.5000	7471B	PPB	29 May 2020 13:46:25	S	Std
10.0 - 1	510	56011	-	10.0000	7471B	PPB	29 May 2020 13:48:42	S	Std
ICV64 - 1	ICV64	21431	3.8431	-	7471B	PPB	29 May 2020 13:51:26	U	SMPL
ICB64 - 1	ICB64	-60	-0.0000	-	7471B	PPB	29 May 2020 13:53:41	U	SMPL
CCV27 - 1	CCV27	27299	4.8925	-	7471B	PPB	29 May 2020 13:55:56	U	SMPL
CCB27 - 1	CCB27	-60	-0.0000	-	7471B	PPB	29 May 2020 13:58:11	U	SMPL
CRA - 1	CRA	1003	0.1900	-	7471B	PPB	29 May 2020 14:00:27	U	SMPL
HighStd - 1	HighStd	54368	9.7332	-	7471B	PPB	29 May 2020 14:02:42	U	SMPL
ChkStd - 1	ChkStd	37104	6.6459	-	7471B	PPB	29 May 2020 14:04:58	U	SMPL
PB129176BL - 1	PBS	-161	-0.0181	-	7471B	PPB	29 May 2020 14:07:15	U	SMPL
PB129176BS - 1	LCSS	18916	3.3934	-	7471B	PPB	29 May 2020 14:09:32	U	SMPL
L2796-01 - 1	CON-WEST-A	515	0.1028	-	7471B	PPB	29 May 2020 14:11:48	U	SMPL
L2796-02 - 1	CON-WEST-B	1885	0.3478	-	7471B	PPB	29 May 2020 14:14:04	U	SMPL
L2796-03 - 1	CON-EAST-A	3786	0.6877	-	7471B	PPB	29 May 2020 14:16:22	U	SMPL
L2796-04 - 1	CON-EAST-B	4481	0.8120	-	7471B	PPB	29 May 2020 14:18:38	U	SMPL
L2796-05 - 1	CON-EAST-C	7809	1.4071	-	7471B	PPB	29 May 2020 14:20:54	U	SMPL
CCV28 - 1	CCV28	25881	4.6389	-	7471B	PPB	29 May 2020 14:23:09	U	SMPL
CCB28 - 1	CCB28	-59	0.0001	-	7471B	PPB	29 May 2020 14:25:24	U	SMPL
L2797-01 - 1	TEST-HOLE-5-COMP	1252	0.2346	-	7471B	PPB	29 May 2020 14:27:40	U	SMPL
L2797-06 - 1	TEST-HOLE-3-COMP	720	0.1394	-	7471B	PPB	29 May 2020 14:29:55	U	SMPL
L2798-01 - 1	NM2-COMP-2020-0-6	10297	1.8521	-	7471B	PPB	29 May 2020 14:32:10	U	SMPL
L2798-01DUP - 1	NM2-COMP-2020-0-6DUP	10013	1.8013	-	7471B	PPB	29 May 2020 14:34:26	U	SMPL
L2798-01MS - 1	NM2-COMP-2020-0-6MS	30582	5.4796	-	7471B	PPB	29 May 2020 14:36:42	U	SMPL
L2798-01MSD - 1	NM2-COMP-2020-0-6MSD	30605	5.4837	-	7471B	PPB	29 May 2020 14:38:58	U	SMPL
L2798-02 - 1	NM45-COMP-2020-0-6	10746	1.9324	-	7471B	PPB	29 May 2020 14:41:15	U	SMPL
L2798-03 - 1	NM87-COMP-2020-0-6	10085	1.8142	-	7471B	PPB	29 May 2020 14:43:33	U	SMPL
L2798-04 - 1	NM88-COMP-2020-0-6	10603	1.9068	-	7471B	PPB	29 May 2020 14:45:49	U	SMPL
L2798-05 - 1	NM105-COMP-2020-0-6	6295	1.1364	-	7471B	PPB	29 May 2020 14:48:05	U	SMPL
CCV29 - 1	CCV29	26497	4.7491	-	7471B	PPB	29 May 2020 14:50:21	U	SMPL
CCB29 - 1	CCB29	-88	-0.0051	-	7471B	PPB	29 May 2020 14:52:36	U	SMPL
L2798-06 - 1	NM106-COMP-2020-0-6	21533	3.8614	-	7471B	PPB	29 May 2020 14:54:53	U	SMPL
L2798-07 - 1	NM57-COMP-2020-0-6	9409	1.6933	-	7471B	PPB	29 May 2020 14:57:08	U	SMPL
L2798-08 - 1	NM98-COMP-2020-0-6	18398	3.3008	-	7471B	PPB	29 May 2020 14:59:24	U	SMPL
L2798-09 - 1	NM99-COMP-2020-0-6	13523	2.4290	-	7471B	PPB	29 May 2020 15:01:40	U	SMPL
L2798-10 - 1	NM100-COMP-2020-0-6	13444	2.4148	-	7471B	PPB	29 May 2020 15:03:57	U	SMPL
L2798-11 - 1	NM101-COMP-2020-0-6	12058	2.1670	-	7471B	PPB	29 May 2020 15:06:13	U	SMPL
L2798-01LX5 - 1		2065	0.3800	-	7471B	PPB	29 May 2020 15:08:29	U	SMPL
L2798-01A - 1		31943	5.7230	-	7471B	PPB	29 May 2020 15:10:46	U	SMPL
CCV30 - 1	CCV30	26640	4.7747	-	7471B	PPB	29 May 2020 15:15:11	U	SMPL
CCB30 - 1	CCB30	-67	-0.0013	-	7471B	PPB	29 May 2020 15:17:26	U	SMPL