



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109369

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	05/29/2020	12:00
2	CAL2	1	Water	NA	NA	21.1	7.00	05/29/2020	12:02
3	CAL3	1	Water	NA	NA	21.1	10.03	05/29/2020	12:03
4	ICV	1	Water	NA	NA	21.1	6.98	05/29/2020	12:05
5	CCV1	1	Water	NA	NA	20.2	2.01	05/29/2020	12:06
6	L2804-01	1	Water	NA	NA	21.2	4.89	05/29/2020	12:10
7	L2804-01DUP	1	Water	NA	NA	21.3	4.90	05/29/2020	12:11
8	L2804-03	1	Water	NA	NA	18.6	4.61	05/29/2020	12:15
9	CCV2	1	Water	NA	NA	20.1	12.02	05/29/2020	12:16

1B109369

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2804 WorkList ID : 139164 Department : Wet-Chemistry Date : 05-29-2020 11:48:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2020	Water	L2804-01 - <u>17</u>	pH	Cool 4 deg C	PSEG01	P12	32370	05/29/2020	9040C
05/29/2020	Water	L2804-03 - <u>C</u>	pH	Cool 4 deg C	PSEG01	P12	32388	05/29/2020	9040C

Date/Time 05/29/2020 11:50
 Received by: JP
 Relinquished by: JP

Date/Time 05/29/2020 13:00
 Received by: JP
 Relinquished by: JP