



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109376

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	05/29/2020	17:25
2	CAL2	1	Water	NA	NA	20.1	7.00	05/29/2020	17:26
3	CAL3	1	Water	NA	NA	20.1	10.02	05/29/2020	17:27
4	ICV	1	Water	NA	NA	20.2	7.00	05/29/2020	17:30
5	CCV1	1	Water	NA	NA	20.2	2.02	05/29/2020	17:32
6	L2805-02	1	Solid	20.03	20	24.2	6.31	05/29/2020	17:33
7	L2813-06	1	Solid	20.04	20	23.2	8.42	05/29/2020	17:35
8	L2814-06	1	Solid	20.02	20	24.7	5.65	05/29/2020	17:36
9	L2814-06DUP	1	Solid	20.03	20	24.8	5.66	05/29/2020	17:37
10	L2822-01	1	Solid	20.02	20	23.8	5.84	05/29/2020	17:40
11	L2822-02	1	Solid	20.03	20	23.3	5.84	05/29/2020	17:41
12	L2822-03	1	Solid	20.01	20	23.5	5.22	05/29/2020	17:43
13	CCV2	1	Water	NA	NA	20.2	12.02	05/29/2020	17:45

WORKLIST(Hardcopy Internal Chain)

VB109376

WorkList Name : CORROSIONITY L2822 WorkList ID : 139171 Department : Wet-Chemistry Date : 05-29-2020 12:58:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2020	Solid	L2805-02	Corrosivity	Cool 4 deg C	PSEG01	P21	OILY-STONE-DRUM	05/29/2020	9045D
05/29/2020	Solid	L2813-06	Corrosivity	Cool 4 deg C	PSEG05	P21	EO-02-052920	05/29/2020	9045D
05/29/2020	Solid	L2814-06	Corrosivity	Cool 4 deg C	PSEG05	P21	PL-02-052920	05/29/2020	9045D
05/28/2020	Solid	L2822-01	Corrosivity	Cool 4 deg C	WEST04	P12	C0AP6	05/28/2020	9045D
05/28/2020	Solid	L2822-02	Corrosivity	Cool 4 deg C	WEST04	P12	C0AP7	05/28/2020	9045D
05/28/2020	Solid	L2822-03	Corrosivity	Cool 4 deg C	WEST04	P12	C0AP8	05/28/2020	9045D

Date/Time 05/29/2020 16:30
 Received by: Jn
 Relinquished by: KS

Date/Time 05/29/2020 18:30
 Received by: KS
 Relinquished by: Jn