



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109385

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/01/2020	11:40
2	CAL2	1	Water	NA	NA	20.2	7.01	06/01/2020	11:41
3	CAL3	1	Water	NA	NA	20.2	10.03	06/01/2020	11:42
4	ICV	1	Water	NA	NA	20.1	7.00	06/01/2020	11:45
5	CCV1	1	Water	NA	NA	20.1	2.01	06/01/2020	11:46
6	L2831-01	1	Water	NA	NA	16.2	6.12	06/01/2020	11:48
7	L2831-01DUP	1	Water	NA	NA	16.3	6.13	06/01/2020	11:49
8	CCV2	1	Water	NA	NA	20.1	12.02	06/01/2020	11:50

1B109385

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2831 WorkList ID : 139197 Department : Wet-Chemistry Date : 06-01-2020 11:32:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
05/29/2020	Water	L2831-01	☒ Corrosivity	Cool 4 deg C	PARS02	P21	WC-20200529	05/29/2020	9040C

Date/Time 06-01-2020 11:35
Received by: Ja
Relinquished by: Ja

Date/Time 06-01-2020 14:00
Received by: Ja
Relinquished by: Ja