



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB109398

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP84579
Chlorine Calibration std, 0.1ppm	WP84572
Chlorine Calibration std, 0.2ppm	WP84573
Chlorine Calibration std, 0.8ppm	WP84575
Chlorine Calibration std, 0.0ppm	WP84571
Chlorine Calibration std, 1.6ppm	WP84577
Residual Chlorine Calibration and CCV std, 0	WP84574

Intercept: -0.0194

Slope: 1.1537

Regression: 0.999365

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	06/02/2020	12:20
2	CAL2	0.1	1	0.000	0.100	0.100	06/02/2020	12:22
3	CAL3	0.2	1	0.000	0.220	0.220	06/02/2020	12:24
4	CAL4	0.4	1	0.000	0.430	0.430	06/02/2020	12:26
5	CAL5	0.8	1	0.000	0.860	0.860	06/02/2020	12:29
6	CAL6	1.6	1	0.000	1.850	1.850	06/02/2020	12:32



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB109398

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4400	0.4400	0.3770	06/02/2020	12:35
2	ICB				1	0.0000	0.0000	0.0000	-0.0130	06/02/2020	12:36
3	CCV1			0.4	1	0.0000	0.4300	0.4300	0.3690	06/02/2020	12:39
4	CCB1				1	0.0000	0.0000	0.0000	-0.0130	06/02/2020	12:41
5	LB109398BL	50	50		1	0.0000	0.0000	0.0000	-0.0130	06/02/2020	12:43
6	LB109398BS	50	50	0.4	1	0.0000	0.4500	0.4500	0.3860	06/02/2020	12:45
7	L2832-01	50	50		1	0.0000	0.0100	0.0100	-0.0040	06/02/2020	12:47
8	L2832-01DUP	50	50		1	0.0000	0.0100	0.0100	-0.0040	06/02/2020	12:49
9	CCV2			0.4	1	0.0000	0.4300	0.4300	0.3690	06/02/2020	12:51
10	CCB2				1	0.0000	0.0000	0.0000	-0.0130	06/02/2020	12:53