

LB109399


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41781



A= 0.0000e+000

B= 2.4030e-004

C= 3.7499e-002

Rho= 0.9999017

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.043	0.043	21	0.000	21				
0.2	0.200	0.208	0.008	711	0.0 %	711				
2.5	2.500	2.500	-0.000	10246	0.0 %	10246				
5.0	5.000	4.933	-0.067	20371	0.0 %	20371				
7.5	7.500	7.440	-0.060	30804	0.0 %	30804				
10.0	10.000	10.077	0.077	41781	0.0 %	41781				

LB109399 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Jun 2020 10:52:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	02 Jun 2020 11:04:47	S	Std
0.2 - 1	58.2	711	-	0.2000	7470A	PPB	02 Jun 2020 11:06:55	S	Std
2.5 - 1	52.5	10246	-	2.5000	7470A	PPB	02 Jun 2020 11:09:02	S	Std
5.0 - 1	55	20371	-	5.0000	7470A	PPB	02 Jun 2020 11:11:10	S	Std
7.5 - 1	57.5	30804	-	7.5000	7470A	PPB	02 Jun 2020 11:13:17	S	Std
10.0 - 1	510	41781	-	10.0000	7470A	PPB	02 Jun 2020 11:17:51	S	Std
ICV67 - 1	ICV67	14969	3.6345	-	7470A	PPB	02 Jun 2020 11:20:21	U	SMPL
ICB67 - 1	ICB67	-20	0.0327	-	7470A	PPB	02 Jun 2020 11:22:26	U	SMPL
CCV39 - 1	CCV39	20593	4.9859	-	7470A	PPB	02 Jun 2020 11:24:32	U	SMPL
CCB39 - 1	CCB39	-3	0.0368	-	7470A	PPB	02 Jun 2020 11:26:37	U	SMPL
CRA - 1	CRA	702	0.2062	-	7470A	PPB	02 Jun 2020 11:28:43	U	SMPL
HighStd - 1	HighStd	41728	10.0646	-	7470A	PPB	02 Jun 2020 11:30:48	U	SMPL
ChkStd - 1	ChkStd	28354	6.8509	-	7470A	PPB	02 Jun 2020 11:32:55	U	SMPL
PB129220BL - 1	PBW	-58	0.0236	-	7470A	PPB	02 Jun 2020 11:37:23	U	SMPL
PB129220BS - 1	LCSW	14413	3.5009	-	7470A	PPB	02 Jun 2020 11:39:30	U	SMPL
L2805-02 - 1	OILY-STONE-DRUM	260	0.1000	-	7470A	PPB	02 Jun 2020 11:41:37	U	SMPL
L2813-06 - 1	EO-02-052920	441	0.1435	-	7470A	PPB	02 Jun 2020 11:43:44	U	SMPL
L2813-06DUP - 1	EO-02-052920DUP	393	0.1319	-	7470A	PPB	02 Jun 2020 11:45:52	U	SMPL
L2813-06MS - 1	EO-02-052920MS	14298	3.4733	-	7470A	PPB	02 Jun 2020 11:47:59	U	SMPL
L2813-06MSD - 1	EO-02-052920MSD	14960	3.6323	-	7470A	PPB	02 Jun 2020 11:50:05	U	SMPL
CCV40 - 1	CCV40	19932	4.8271	-	7470A	PPB	02 Jun 2020 11:52:11	U	SMPL
CCB40 - 1	CCB40	-125	0.0075	-	7470A	PPB	02 Jun 2020 11:54:17	U	SMPL
L2814-06 - 1	PL-02-052920	376	0.1279	-	7470A	PPB	02 Jun 2020 11:56:23	U	SMPL
PB129221BL - 1	PBW	-59	0.0233	-	7470A	PPB	02 Jun 2020 11:58:29	U	SMPL
PB129221BS - 1	LCSW	15082	3.6617	-	7470A	PPB	02 Jun 2020 12:00:35	U	SMPL
L2796-07 - 1	N48963	354	0.1226	-	7470A	PPB	02 Jun 2020 12:02:42	U	SMPL
L2796-07DUP - 1	N48963DUP	352	0.1221	-	7470A	PPB	02 Jun 2020 12:04:49	U	SMPL
L2796-07MS - 1	N48963MS	17129	4.1535	-	7470A	PPB	02 Jun 2020 12:13:39	U	SMPL
L2796-07MSD - 1	N48963MSD	16826	4.0807	-	7470A	PPB	02 Jun 2020 12:15:45	U	SMPL
PB129234BL - 1	PBW	-131	0.0060	-	7470A	PPB	02 Jun 2020 12:17:53	U	SMPL
PB129234BS - 1	LCSW	14503	3.5225	-	7470A	PPB	02 Jun 2020 12:20:01	U	SMPL
L2795-05 - 1	WATER-COMP-214-215	566	0.1735	-	7470A	PPB	02 Jun 2020 12:22:06	U	SMPL
CCV41 - 1	CCV41	20372	4.9328	-	7470A	PPB	02 Jun 2020 12:24:13	U	SMPL
CCB41 - 1	CCB41	12	0.0404	-	7470A	PPB	02 Jun 2020 12:28:49	U	SMPL
L2795-05DUP - 1	WATER-COMP-214-215DUP	659	0.1959	-	7470A	PPB	02 Jun 2020 12:30:55	U	SMPL
L2795-05MS - 1	WATER-COMP-214-215MS	17826	4.3210	-	7470A	PPB	02 Jun 2020 12:35:49	U	SMPL
L2795-05MSD - 1	WATER-COMP-214-215MSD	17231	4.1781	-	7470A	PPB	02 Jun 2020 12:37:54	U	SMPL
L2831-02 - 1	WC-20200529	2245	0.5770	-	7470A	PPB	02 Jun 2020 12:40:01	U	SMPL
L2832-01 - 1	20200383-DI-WATER	564	0.1730	-	7470A	PPB	02 Jun 2020 12:42:07	U	SMPL
PB129220TB - 1	PB129196TB	-35	0.0291	-	7470A	PPB	02 Jun 2020 12:46:47	U	SMPL
PB129234TB - 1	PB129189TB	-25	0.0315	-	7470A	PPB	02 Jun 2020 12:48:53	U	SMPL
L2813-06LX5 - 1		38	0.0466	-	7470A	PPB	02 Jun 2020 12:51:01	U	SMPL
L2813-06A - 1		17678	4.2855	-	7470A	PPB	02 Jun 2020 12:53:09	U	SMPL
CCV42 - 1	CCV42	20384	4.9357	-	7470A	PPB	02 Jun 2020 12:55:15	U	SMPL
CCB42 - 1	CCB42	-118	0.0091	-	7470A	PPB	02 Jun 2020 12:57:21	U	SMPL
L2796-07LX5 - 1		27	0.0440	-	7470A	PPB	02 Jun 2020 12:59:27	U	SMPL
L2796-07A - 1		17802	4.3153	-	7470A	PPB	02 Jun 2020 13:01:34	U	SMPL
L2795-05LX5 - 1		-5	0.0363	-	7470A	PPB	02 Jun 2020 13:03:40	U	SMPL
L2795-05A - 1		18276	4.4292	-	7470A	PPB	02 Jun 2020 13:05:47	U	SMPL
CCV43 - 1	CCV43	20171	4.8845	-	7470A	PPB	02 Jun 2020 13:07:53	U	SMPL
CCB43 - 1	CCB43	-105	0.0123	-	7470A	PPB	02 Jun 2020 13:10:00	U	SMPL

03/29/21/73
6/19/20