

LB109404


245.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38495



A= 0.0000e+000

B= 2.6260e-004

C= -2.0411e-002

Rho= 0.9998234

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	-54	0.000	-54				
0.05	0.050	0.038	-0.012	223	0.0 %	223				
0.2	0.200	0.256	0.056	1051	0.0 %	1051				
2.5	2.500	2.533	0.033	9722	0.0 %	9722				
5.0	5.000	5.014	0.014	19170	0.0 %	19170				
7.5	7.500	7.356	-0.144	28092	0.0 %	28092				
10.0	10.000	10.088	0.088	38495	0.0 %	38495				

LB109404 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 02 Jun 2020 14:09:45

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	SD	-54	-	0.0000245.1	PPB	02 Jun 2020 14:18:54	S	Std
0.05 - 1	SD 0.05	223	-	0.0500245.1	PPB	02 Jun 2020 14:23:44	S	Std
0.2 - 1	SD 0.2	1051	-	0.2000245.1	PPB	02 Jun 2020 14:26:01	S	Std
2.5 - 1	SD 2.5	9722	-	2.5000245.1	PPB	02 Jun 2020 14:28:17	S	Std
5.0 - 1	SD 5.0	19170	-	5.0000245.1	PPB	02 Jun 2020 14:30:34	S	Std
7.5 - 1	SD 7.5	28092	-	7.5000245.1	PPB	02 Jun 2020 14:32:49	S	Std
10.0 - 1	SD 10.0	38495	-	10.0000245.1	PPB	02 Jun 2020 14:41:21	S	Std
ICV68 - 1	ICV68	15266	3.9884	-245.1	PPB	02 Jun 2020 14:44:31	U	SMPL
ICB68 - 1	ICB68	72	-0.0015	-245.1	PPB	02 Jun 2020 14:46:46	U	SMPL
CCV44 - 1	CCV44	19685	5.1488	-245.1	PPB	02 Jun 2020 14:49:01	U	SMPL
CCB44 - 1	CCB44	60	-0.0047	-245.1	PPB	02 Jun 2020 14:51:16	U	SMPL
CRA - 1	CRA	241	0.0429	-245.1	PPB	02 Jun 2020 14:53:31	U	SMPL
HighStd - 1	HighStd	38600	10.1158	-245.1	PPB	02 Jun 2020 14:55:46	U	SMPL
ChkStd - 1	ChkStd	26857	7.0322	-245.1	PPB	02 Jun 2020 14:58:01	U	SMPL
PB129237BL - 1	PBW	-47	-0.0328	-245.1	PPB	02 Jun 2020 15:02:50	U	SMPL
PB129237BS - 1	LCSW	15577	4.0701	-245.1	PPB	02 Jun 2020 15:07:42	U	SMPL
L2494-02 - 1	20200382-MANASQUAN-WATER	798	0.1891	-245.1	PPB	02 Jun 2020 15:09:58	U	SMPL
L2494-02DUP - 1	20200382-MANASQUAN-WATERDUP	779	0.1842	-245.1	PPB	02 Jun 2020 15:15:10	U	SMPL
L2494-02MS - 1	20200382-MANASQUAN-WATERMS	16101	4.2077	-245.1	PPB	02 Jun 2020 15:17:27	U	SMPL
L2494-02MSD - 1	20200382-MANASQUAN-WATERMSD	16198	4.2331	-245.1	PPB	02 Jun 2020 15:19:44	U	SMPL
CCV45 - 1	CCV45	19376	5.0677	-245.1	PPB	02 Jun 2020 15:21:59	U	SMPL
CCB45 - 1	CCB45	15	-0.0165	-245.1	PPB	02 Jun 2020 15:24:14	U	SMPL
L2494-02LX5 - 1		167	0.0234	-245.1	PPB	02 Jun 2020 15:26:31	U	SMPL
L2494-02A - 1		16242	4.2447	-245.1	PPB	02 Jun 2020 15:28:46	U	SMPL
CCV46 - 1	CCV46	19207	5.0233	-245.1	PPB	02 Jun 2020 15:31:01	U	SMPL
CCB46 - 1	CCB46	-148	-0.0593	-245.1	PPB	02 Jun 2020 15:33:16	U	SMPL