



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109406

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/02/2020	10:05
2	L2804-01		1	8	100.00	>212.0	>212.0	06/02/2020	10:40
3	L2804-01DUP		1	8	100.00	>212.0	>212.0	06/02/2020	11:10
4	L2804-03		1	9	100.00	>212.0	>212.0	06/02/2020	11:45
5	L2831-01		1	8	100.00	>212.0	>212.0	06/02/2020	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp6120

WorkList ID : 139188

Department : Wet-Chemistry

Date : 06-01-2020 09:07:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2020	Water	L2804-01	Flash Point	Cool 4 deg C	PSEG01	P12	32370	05/29/2020	1010A
06/05/2020	Water	L2804-03	Flash Point	Cool 4 deg C	PSEG01	P12	32388	05/29/2020	1010A
06/05/2020	Water	L2831-01	Flash Point	Cool 4 deg C	PARS02	P21	WC-20200529	05/29/2020	1010A

Date/Time 06/02/2020 09:30

Received by: KS

Relinquished by: DK

Date/Time 06/02/2020 12:25

Received by: DK

Relinquished by: KS