



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 6/4/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:00
In Date: 06/03/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/04/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109408

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L2814-07	PL-01-060320	17	1.12	9.68	8.91	91.0	
L2814-08	PL-01-060320	18	1.13	9.88	8.41	83.2	
L2814-09	PL-01-060320-1	19	1.18	9.59	9.18	95.1	
L2814-10	PL-01-060320-2	20	1.13	9.7	8.95	91.2	
L2814-11	PL-01-060320-3	21	1.1	9.91	9.02	89.9	
L2814-12	PL-01-060320-4	22	1.12	9.55	8.89	92.2	
L2837-04	SP-1	1	1.13	9.59	9.36	97.3	
L2838-01	VH-DUCT-A	2	1.14	9.55	8.88	92.0	
L2838-03	VH-DUCT-B	3	1.16	9.86	8.8	87.8	
L2840-01	DRUM-OIL	4	1.12	9.64	8.74	89.4	
L2840-03	DRUM-OIL	5	1.19	9.66	8.78	89.6	
L2854-01	B-1	6	1.19	9.8	8.47	84.6	
L2854-02	B-2	7	1.12	9.76	8.24	82.4	
L2854-03	B-3	8	1.14	9.82	8.37	83.3	
L2854-04	B-4	9	1.15	9.96	9.00	89.1	
L2854-05	B-5	10	1.13	9.97	8.76	86.3	
L2858-01	CHRT27247	23	1.15	9.96	9.03	89.4	
L2870-01	PCB-060120-SR-01	11	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2870-03	PCB-060120-SR-02	12	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2870-04	PCB-060120-SR-03	13	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2871-01	WARREN-TP	14	1.16	9.52	8.63	89.4	
L2871-02	WARREN-TP-A	15	1.19	9.7	8.76	89.0	
L2871-03	WARREN-TP-B	16	1.19	9.77	8.86	89.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$