



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109410

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/03/2020	08:20
2	CAL2	1	Water	NA	NA	20.1	7.00	06/03/2020	08:21
3	CAL3	1	Water	NA	NA	20.1	10.03	06/03/2020	08:22
4	ICV	1	Water	NA	NA	20.1	6.99	06/03/2020	08:24
5	CCV1	1	Water	NA	NA	20.2	2.01	06/03/2020	08:25
6	L2840-04	1	Water	NA	NA	21.3	5.37	06/03/2020	08:26
7	L2840-04DUP	1	Water	NA	NA	21.4	5.38	06/03/2020	08:28
8	CCV2	1	Water	NA	NA	20.1	12.02	06/03/2020	08:40

LB109410

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2840 WorkList ID : 139243 Department : Wet-Chemistry Date : 06-03-2020 08:13:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/02/2020	Water	L2840-04	pH	Cool 4 deg C	PSEG01	P41	TOTE-OIL	06/02/2020	9040C

Date/Time 06/03/2020 8:15
Received by: JS
Relinquished by: JS

Date/Time 06/03/2020 9:30
Received by: JS
Relinquished by: JS