



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109415

Slope : 97.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	06/03/2020	13:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/03/2020	13:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/03/2020	13:02
4	ICV	1	Water	NA	NA	20.1	6.99	06/03/2020	13:05
5	CCV1	1	Water	NA	NA	20.2	2.01	06/03/2020	13:06
6	L2857-01	1	Water	NA	NA	15.6	6.58	06/03/2020	13:10
7	L2857-01DUP	1	Water	NA	NA	15.7	6.59	06/03/2020	13:11
8	CCV2	1	Water	NA	NA	20.2	12.02	06/03/2020	13:15

VB 109415

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2857 WorkList ID : 139266 Department : Wet-Chemistry Date : 06-03-2020 12:42:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/02/2020	Water	L2857-01	✓ pH	Cool 4 deg C	APT101	O11	WGGM-20-WWW-TSF-019	06/02/2020	SM 4500-H B

Date/Time 06/03/2020 12:50
Received by: JQ
Relinquished by: JQ

Date/Time 06/03/2020 14:00
Received by: JQ
Relinquished by: JQ