



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109421

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/03/2020	17:05
2	CAL2	1	Water	NA	NA	20.2	7.01	06/03/2020	17:06
3	CAL3	1	Water	NA	NA	20.2	10.03	06/03/2020	17:07
4	ICV	1	Water	NA	NA	20.2	7.01	06/03/2020	17:09
5	CCV1	1	Water	NA	NA	20.1	2.01	06/03/2020	17:10
6	L2814-13	1	Solid	20.03	20	20.8	7.35	06/03/2020	17:11
7	L2814-13DUP	1	Solid	20.04	20	20.9	7.36	06/03/2020	17:12
8	L2837-02	1	Solid	20.02	20	20.7	6.01	06/03/2020	17:15
9	L2871-04	1	Solid	20.03	20	20.6	6.78	06/03/2020	17:19
10	CCV2	1	Water	NA	NA	20.1	12.02	06/03/2020	17:20

13109421

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2814 WorkList ID : 139270 Department : Wet-Chemistry Date : 06-03-2020 13:44:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/03/2020	Solid	L2814-13	Corrosivity	Cool 4 deg C	PSEG05		PL-01-060320	06/03/2020	9045D
06/02/2020	Solid	L2837-02	Corrosivity	Cool 4 deg C	PSEG01	P31	SP	06/02/2020	9045D
06/03/2020	Solid	L2871-04	Corrosivity	Cool 4 deg C	PSEG01	O12	WARREN-TP	06/03/2020	9045D

Date/Time 06/03/2020 13:50
Received by: JA
Relinquished by: KS

Date/Time 06/03/2020 18:00
Received by: KS
Relinquished by: JD