



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109422

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/03/2020	17:05
2	CAL2	1	Water	NA	NA	20.2	7.01	06/03/2020	17:06
3	CAL3	1	Water	NA	NA	20.2	10.03	06/03/2020	17:07
4	ICV	1	Water	NA	NA	20.2	7.01	06/03/2020	17:09
5	CCV1	1	Water	NA	NA	20.2	2.01	06/03/2020	17:25
6	L2858-03	1	Water	NA	NA	20.4	6.94	06/03/2020	17:26
7	L2858-03DUP	1	Water	NA	NA	20.3	6.95	06/03/2020	17:27
8	L2859-02	1	Water	NA	NA	20.2	6.45	06/03/2020	17:28
9	CCV2	1	Water	NA	NA	20.2	12.02	06/03/2020	17:30

VB109422

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2859 WorkList ID : 139275 Department : Wet-Chemistry Date : 06-03-2020 16:27:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/03/2020	Water	L2858-03 - <i>CS</i>	Corrosivity	Cool 4 deg C	PSEG01	O13	N28427	06/03/2020	9040C
06/03/2020	Water	L2859-02 - <i>CS</i>	Corrosivity	Cool 4 deg C	PSEG01	O13	SP-UST	06/03/2020	9040C

Date/Time 06/03/2020 16:30
 Received by: *JS*
 Relinquished by: _____

Date/Time 06/03/2020 17:40
 Received by: *KS*
 Relinquished by: *JS*