



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB109425
Analysis Date: 06/04/2020
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 06/04/2020
Extraction IN Time: 07:50
Extraction OUT Time: 08:35

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB109425BL	LB109425BL	WATER	1.3	1000	100	2.8964	2.8964	3.02	2.8964	2.8964	0.0000	0
2	LB109425BS	LB109425BS	WATER	1.3	1000	100	2.1303	2.1303	3.01	2.1494	2.1494	0.0191	19.1
3	L2493-03	MW-03	WATER	1.3	1000	100	2.9701	2.9701	3.03	2.9714	2.9714	0.0013	1.3
4	L2493-04	L2493-03MS	WATER	1.3	1000	100	2.8547	2.8547	3.04	2.8760	2.8760	0.0213	21.3
5	L2493-05	L2493-03MSD	WATER	1.3	1000	100	2.7411	2.7411	3.03	2.7621	2.7621	0.0210	21
6	L2493-06	MW-02	WATER	1.3	1000	100	2.9889	2.9889	3.02	3.0046	3.0046	0.0157	15.7
7	L2858-02	N28427	WATER	1.3	1000	100	2.9580	2.9580	3.04	2.9615	2.9615	0.0035	3.5
8	L2859-01	SP-UST	WATER	1.6	1000	100	2.9671	2.9671	3.05	2.9685	2.9685	0.0014	1.4
9	L2883-02	GRAB	WATER	1.6	1000	100	2.9650	2.9650	3.03	2.9884	2.9884	0.0234	23.4



QC Batch# LB109425 Test: TPH Analysis Date: 06/04/2020

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2693
PH STRIPS BOX-0.2.5	M4462
Sodium Sulfate	EP2046
1:1 HCL	WP80893
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP80894
LCSWD	NA	NA
MS/MSD	10.00 ML	WP80895

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>10:02</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time1:	<u>09:00</u>		
Bal Check Time:	<u>08:10</u>	Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>10:35</u>
		Out Time1	<u>10:00</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>11:46</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time2:	<u>11:10</u>		
Bal Check Time:	<u>13:00</u>	Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>12:15</u>
		Out Time2:	<u>11:45</u>		

109625

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH L2883

WorkList ID : 139283

Department : Wet-Chemistry

Date : 06-04-2020 07:38:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2020	Water	L2493-03 <u>B</u>	TPH	Conc H2SO4 tc	DVIR01	P11	MW-03	05/29/2020	1664A
06/15/2020	Water	L2493-04 <u>✓</u>	TPH	Conc H2SO4 tc	DVIR01	P11	L2493-MS	05/29/2020	1664A
06/15/2020	Water	L2493-05 <u>✓</u>	TPH	Conc H2SO4 tc	DVIR01	P11	L2493-MSD	05/29/2020	1664A
06/15/2020	Water	L2493-06 <u>B</u>	TPH	Conc H2SO4 tc	DVIR01	P11	MW-02	05/29/2020	1664A
06/08/2020	Water	L2858-02 - <u>✓</u>	TPH	Conc H2SO4 tc	PSEG01	O13	N28427	06/03/2020	1664A
06/08/2020	Water	L2859-01 - <u>✓</u>	TPH	Conc H2SO4 tc	PSEG01	O13	SP-UST	06/03/2020	1664A
06/16/2020	Water	L2883-02 <u>✓</u>	TPH	Conc H2SO4 tc	ARAM01	O31	GRAB	06/03/2020	1664A

Date/Time 06/04/2020 7:40
 Received by: Jr
 Relinquished by: Jr

Date/Time 06/04/2020 10:13
 Received by: Jr
 Relinquished by: Jr