

LB109426


7470A

INSTRUMENT ID: CVL

Linear

μ Abs.:

36579



A= 0.0000e+000

B= 2.7241e-004

C= -8.7381e-003

Rho= 0.9998403

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.020	-0.020	-40	0.000	-40				
0.2	0.200	0.185	-0.015	710	0.0 %	710				
2.5	2.500	2.462	-0.038	9070	0.0 %	9070				
5.0	5.000	5.146	0.146	18922	0.0 %	18922				
7.5	7.500	7.471	-0.029	27459	0.0 %	27459				
10.0	10.000	9.956	-0.044	36579	0.0 %	36579				

LB109426 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Jun 2020 10:04:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	-40	-	0.0000	7470A	PPB	04 Jun 2020 10:12:54	S	Std
0.2 - 1	50.2	710	-	0.2000	7470A	PPB	04 Jun 2020 10:15:01	S	Std
2.5 - 1	52.5	9070	-	2.5000	7470A	PPB	04 Jun 2020 10:17:09	S	Std
5.0 - 1	55	18922	-	5.0000	7470A	PPB	04 Jun 2020 10:19:17	S	Std
7.5 - 1	57.5	27459	-	7.5000	7470A	PPB	04 Jun 2020 10:21:24	S	Std
10.0 - 1	510	36579	-	10.0000	7470A	PPB	04 Jun 2020 10:28:25	S	Std
ICV70 - 1	ICV70	13540	3.6797	-7470A	PPB	PPB	04 Jun 2020 10:30:55	U	SMPL
ICB70 - 1	ICB70	-173	-0.0559	-7470A	PPB	PPB	04 Jun 2020 10:33:00	U	SMPL
CCV55 - 1	CCV55	19088	5.1910	-7470A	PPB	PPB	04 Jun 2020 10:35:06	U	SMPL
CCB55 - 1	CCB55	-153	-0.0504	-7470A	PPB	PPB	04 Jun 2020 10:37:12	U	SMPL
CRA - 1	CRA	704	0.1830	-7470A	PPB	PPB	04 Jun 2020 10:39:18	U	SMPL
HighStd - 1	HighStd	36644	9.9735	-7470A	PPB	PPB	04 Jun 2020 10:41:24	U	SMPL
ChkStd - 1	ChkStd	25072	6.8211	-7470A	PPB	PPB	04 Jun 2020 10:43:29	U	SMPL
PB129276BL - 1	PBW	-180	-0.0578	-7470A	PPB	PPB	04 Jun 2020 10:45:37	U	SMPL
PB129276BS - 1	LCSW	13591	3.6936	-7470A	PPB	PPB	04 Jun 2020 10:47:44	U	SMPL
L2837-02 - 1	SP	-134	-0.0452	-7470A	PPB	PPB	04 Jun 2020 10:49:51	U	SMPL
L2837-02DUP - 1	SPDUP	-79	-0.0303	-7470A	PPB	PPB	04 Jun 2020 10:51:58	U	SMPL
L2837-02MS - 1	SPMS	12963	3.5225	-7470A	PPB	PPB	04 Jun 2020 10:54:06	U	SMPL
L2837-02MSD - 1	SPMSD	13691	3.7208	-7470A	PPB	PPB	04 Jun 2020 10:56:14	U	SMPL
L2838-02 - 1	VH-DUCT-A	-136	-0.0458	-7470A	PPB	PPB	04 Jun 2020 10:58:20	U	SMPL
CCV56 - 1	CCV56	19180	5.2161	-7470A	PPB	PPB	04 Jun 2020 11:00:26	U	SMPL
CCB56 - 1	CCB56	-161	-0.0526	-7470A	PPB	PPB	04 Jun 2020 11:02:32	U	SMPL
L2838-04 - 1	VH-DUCT-B	-49	-0.0221	-7470A	PPB	PPB	04 Jun 2020 11:04:38	U	SMPL
L2840-01 - 1	DRUM-OIL	-56	-0.0240	-7470A	PPB	PPB	04 Jun 2020 11:06:44	U	SMPL
PB129277BL - 1	PBW	-62	-0.0256	-7470A	PPB	PPB	04 Jun 2020 11:08:50	U	SMPL
PB129277BS - 1	LCSW	13507	3.6707	-7470A	PPB	PPB	04 Jun 2020 11:13:26	U	SMPL
L2840-04 - 1	TOTE-OIL	-9	-0.0112	-7470A	PPB	PPB	04 Jun 2020 11:15:32	U	SMPL
L2840-04DUP - 1	TOTE-OILDUP	74	0.0114	-7470A	PPB	PPB	04 Jun 2020 11:17:39	U	SMPL
L2840-04MS - 1	TOTE-OILMS	16358	4.4473	-7470A	PPB	PPB	04 Jun 2020 11:22:38	U	SMPL
L2840-04MSD - 1	TOTE-OILMSD	16354	4.4463	-7470A	PPB	PPB	04 Jun 2020 11:24:45	U	SMPL
CCV57 - 1	CCV57	18979	5.1613	-7470A	PPB	PPB	04 Jun 2020 11:26:53	U	SMPL
CCB57 - 1	CCB57	-150	-0.0496	-7470A	PPB	PPB	04 Jun 2020 11:28:59	U	SMPL
PB129276TB - 1	PB129261TB	-46	-0.0213	-7470A	PPB	PPB	04 Jun 2020 11:31:05	U	SMPL
PB129277TB - 1	PB129272TB	-60	-0.0251	-7470A	PPB	PPB	04 Jun 2020 11:33:12	U	SMPL
L2837-02LX5 - 1		-41	-0.0199	-7470A	PPB	PPB	04 Jun 2020 11:35:18	U	SMPL
L2837-02A - 1		14069	3.8238	-7470A	PPB	PPB	04 Jun 2020 11:37:24	U	SMPL
L2840-04LX5 - 1		-69	-0.0275	-7470A	PPB	PPB	04 Jun 2020 11:44:25	U	SMPL
L2840-04A - 1		14717	4.0003	-7470A	PPB	PPB	04 Jun 2020 11:46:31	U	SMPL
CCV58 - 1	CCV58	18520	5.0363	-7470A	PPB	PPB	04 Jun 2020 11:48:38	U	SMPL
CCB58 - 1	CCB58	-175	-0.0564	-7470A	PPB	PPB	04 Jun 2020 11:50:45	U	SMPL