



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109436

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/04/2020	10:15
2	L2859-02		1	8	100.00	>212.0	>212.0	06/04/2020	10:50
3	L2859-02DUP		1	8	100.00	>212.0	>212.0	06/04/2020	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB109436

WorkList Name : fp6420 WorkList ID : 139306 Department : Wet-Chemistry Date : 06-04-2020 09:33:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2020	Water	L2859-02	C Flash Point	Cool 4 deg C	PSEG01	O13	SP-UST	06/03/2020	1010A

Date/Time 06/04/2020 09:45
 Received by: KS
 Relinquished by: JS

Date/Time 06/04/2020 11:25
 Received by: JS
 Relinquished by: KS