



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109442

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/04/2020	17:00
2	CAL2	1	Water	NA	NA	20.2	7.00	06/04/2020	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	06/04/2020	17:02
4	ICV	1	Water	NA	NA	20.2	7.00	06/04/2020	17:05
5	CCV1	1	Water	NA	NA	20.1	2.01	06/04/2020	17:06
6	L2889-01	1	Water	NA	NA	22.8	7.47	06/04/2020	17:10
7	L2889-01DUP	1	Water	NA	NA	22.9	7.48	06/04/2020	17:11
8	CCV2	1	Water	NA	NA	20.2	12.02	06/04/2020	17:14

LB109442

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2889 WorkList ID : 139320 Department : Wet-Chemistry Date : 06-04-2020 16:34:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2889-01	-C pH	Cool 4 deg C	PSEG01	O23	FDW-115	06/04/2020	9040C

Date/Time 06/04/2020 16:35
Received by: JP
Relinquished by: JP

Date/Time 06/04/2020 17:20
Received by: JP
Relinquished by: JP