

LB109447


7470A

INSTRUMENT 1D: CV1

Linear

 μ Abs.:

39260



A= 0.0000e+000

B= 2.5186e-004

C= -3.0671e-002

Rho= 0.9995605

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.023	-0.023	32	0.000	32				
0.2	0.200	0.173	-0.027	809	0.0 %	809				
2.5	2.500	2.549	0.049	10242	0.0 %	10242				
5.0	5.000	4.935	-0.065	19715	0.0 %	19715				
7.5	7.500	7.709	0.209	30729	0.0 %	30729				
10.0	10.000	9.857	-0.143	39260	0.0 %	39260				

LB109447 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 05 Jun 2020 10:01:44					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	32	-	0.0000	7470A	PPB	05 Jun 2020 10:06:31	S	Std
0.2 - 1	50.2	809	-	0.2000	7470A	PPB	05 Jun 2020 10:08:39	S	Std
2.5 - 1	52.5	10242	-	2.5000	7470A	PPB	05 Jun 2020 10:10:47	S	Std
5.0 - 1	55	19715	-	5.0000	7470A	PPB	05 Jun 2020 10:12:54	S	Std
7.5 - 1	57.5	30729	-	7.5000	7470A	PPB	05 Jun 2020 10:17:32	S	Std
10.0 - 1	510	39260	-	10.0000	7470A	PPB	05 Jun 2020 10:19:38	S	Std
ICV72 - 1	ICV72	14954	3.7356	-	7470A	PPB	05 Jun 2020 10:22:12	U	SMPL
ICB72 - 1	ICB72	-68	-0.0478	-	7470A	PPB	05 Jun 2020 10:24:18	U	SMPL
CCV62 - 1	CCV62	19570	4.8982	-	7470A	PPB	05 Jun 2020 10:26:24	U	SMPL
CCB62 - 1	CCB62	-74	-0.0493	-	7470A	PPB	05 Jun 2020 10:28:30	U	SMPL
CRA - 1	CRA	804	0.1718	-	7470A	PPB	05 Jun 2020 10:30:35	U	SMPL
HighStd - 1	HighStd	39192	9.8402	-	7470A	PPB	05 Jun 2020 10:32:42	U	SMPL
ChkStd - 1	ChkStd	27380	6.8652	-	7470A	PPB	05 Jun 2020 10:34:47	U	SMPL
PB129312BL - 1	PBW	-167	-0.0727	-	7470A	PPB	05 Jun 2020 10:36:55	U	SMPL
PB129312BS - 1	LCSW	15887	3.9706	-	7470A	PPB	05 Jun 2020 10:39:04	U	SMPL
L2814-13 - 1	PL-01-060320	-110	-0.0584	-	7470A	PPB	05 Jun 2020 10:41:11	U	SMPL
L2814-13DUP - 1	PL-01-060320DUP	-35	-0.0395	-	7470A	PPB	05 Jun 2020 10:43:18	U	SMPL
L2814-13MS - 1	PL-01-060320MS	15048	3.7593	-	7470A	PPB	05 Jun 2020 10:45:25	U	SMPL
L2814-13MSD - 1	PL-01-060320MSD	15023	3.7530	-	7470A	PPB	05 Jun 2020 10:47:33	U	SMPL
L2871-04 - 1	WARREN-TP	-98	-0.0554	-	7470A	PPB	05 Jun 2020 10:49:39	U	SMPL
CCV63 - 1	CCV63	19222	4.8105	-	7470A	PPB	05 Jun 2020 10:51:45	U	SMPL
CCB63 - 1	CCB63	-70	-0.0483	-	7470A	PPB	05 Jun 2020 10:53:50	U	SMPL
PB129313BL - 1	PBW	-36	-0.0397	-	7470A	PPB	05 Jun 2020 10:55:56	U	SMPL
PB129313BS - 1	LCSW	15241	3.8079	-	7470A	PPB	05 Jun 2020 10:58:02	U	SMPL
L2858-03 - 1	N28427	-71	-0.0486	-	7470A	PPB	05 Jun 2020 11:00:09	U	SMPL
L2858-03DUP - 1	N28427DUP	0	-0.0307	-	7470A	PPB	05 Jun 2020 11:02:16	U	SMPL
L2858-03MS - 1	N28427MS	15905	3.9751	-	7470A	PPB	05 Jun 2020 11:04:22	U	SMPL
L2858-03MSD - 1	N28427MSD	14718	3.6762	-	7470A	PPB	05 Jun 2020 11:06:29	U	SMPL
L2859-02 - 1	SP-UST	-98	-0.0554	-	7470A	PPB	05 Jun 2020 11:08:36	U	SMPL
PB129314BL - 1	PBW	-9	-0.0329	-	7470A	PPB	05 Jun 2020 11:10:44	U	SMPL
PB129314BS - 1	LCSW	14860	3.7119	-	7470A	PPB	05 Jun 2020 11:12:52	U	SMPL
L2839-25 - 1	NM-EB-09	-157	-0.0702	-	7470A	PPB	05 Jun 2020 11:14:59	U	SMPL
CCV64 - 1	CCV64	19533	4.8889	-	7470A	PPB	05 Jun 2020 11:17:05	U	SMPL
CCB64 - 1	CCB64	-75	-0.0496	-	7470A	PPB	05 Jun 2020 11:19:11	U	SMPL
L2839-26 - 1	NM-EB-10	-113	-0.0591	-	7470A	PPB	05 Jun 2020 11:21:17	U	SMPL
L2858-02 - 1	N28427	1221	0.2768	-	7470A	PPB	05 Jun 2020 11:23:23	U	SMPL
L2858-02DUP - 1	N28427DUP	1305	0.2980	-	7470A	PPB	05 Jun 2020 11:25:30	U	SMPL
L2858-02MS - 1	N28427MS	14693	3.6699	-	7470A	PPB	05 Jun 2020 11:27:36	U	SMPL
L2858-02MSD - 1	N28427MSD	14508	3.6233	-	7470A	PPB	05 Jun 2020 11:29:43	U	SMPL
L2859-01 - 1	SP-UST	44	-0.0196	-	7470A	PPB	05 Jun 2020 11:31:50	U	SMPL
CCV65 - 1	CCV65	19388	4.8524	-	7470A	PPB	05 Jun 2020 11:33:57	U	SMPL
CCB65 - 1	CCB65	-45	-0.0420	-	7470A	PPB	05 Jun 2020 11:36:04	U	SMPL
PB129312TB - 1	PB129283TB	-14	-0.0342	-	7470A	PPB	05 Jun 2020 11:38:12	U	SMPL
PB129313TB - 1	PB129288TB	-25	-0.0370	-	7470A	PPB	05 Jun 2020 11:40:19	U	SMPL
L2814-13LX5 - 1		-13	-0.0339	-	7470A	PPB	05 Jun 2020 11:42:25	U	SMPL
L2814-13A - 1		14903	3.7228	-	7470A	PPB	05 Jun 2020 11:44:32	U	SMPL
L2858-03LX5 - 1		-88	-0.0528	-	7470A	PPB	05 Jun 2020 11:46:38	U	SMPL
L2858-03A - 1		16198	4.0489	-	7470A	PPB	05 Jun 2020 11:48:45	U	SMPL
L2858-02LX5 - 1		144	0.0056	-	7470A	PPB	05 Jun 2020 11:50:51	U	SMPL
L2858-02A - 1		17469	4.3690	-	7470A	PPB	05 Jun 2020 11:52:57	U	SMPL
CCV67 - 1	CCV67	19168	4.7969	-	7470A	PPB	05 Jun 2020 11:55:04	U	SMPL
CCB67 - 1	CCB67	-101	-0.0561	-	7470A	PPB	05 Jun 2020 11:57:11	U	SMPL