



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109456

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/05/2020	13:00
2	CAL2	1	Water	NA	NA	20.1	7.00	06/05/2020	13:01
3	CAL3	1	Water	NA	NA	20.1	10.03	06/05/2020	13:02
4	ICV	1	Water	NA	NA	20.1	7.01	06/05/2020	13:05
5	CCV1	1	Water	NA	NA	20.2	2.01	06/05/2020	13:06
6	L2926-01	1	Water	NA	NA	16.4	6.46	06/05/2020	13:10
7	L2926-01DUP	1	Water	NA	NA	16.5	6.47	06/05/2020	13:11
8	CCV2	1	Water	NA	NA	20.1	12.02	06/05/2020	13:12

VB109456

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L2926 WorkList ID : 139365 Department : Wet-Chemistry Date : 06-05-2020 12:42:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2926-01 -D	Corrosivity	Cool 4 deg C	TETR06	O41	TANK-1-20200604	06/04/2020	9040C

Date/Time 06/05/2020 12:56
 Received by: JP
 Relinquished by: JP

Date/Time 06/05/2020 16:10
 Received by: JP
 Relinquished by: JP