


LB109460

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.025	-0.025	-19	0.000	-19				
0.2	0.200	0.184	-0.016	772	0.0 %	772				
2.5	2.500	2.658	0.158	10149	0.0 %	10149				
5.0	5.000	4.882	-0.118	18573	0.0 %	18573				
7.5	7.500	7.445	-0.055	28284	0.0 %	28284				
10.0	10.000	10.056	0.056	38178	0.0 %	38178				

LB109460 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Jun 2020 13:40:53

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-19	-	0.0000	7470A	PPB	05 Jun 2020 13:42:04	S	Std
0.2 - 1	50.2	772	-	0.2000	7470A	PPB	05 Jun 2020 13:44:12	S	Std
2.5 - 1	52.5	10149	-	2.5000	7470A	PPB	05 Jun 2020 13:46:20	S	Std
5.0 - 1	55	18573	-	5.0000	7470A	PPB	05 Jun 2020 13:48:27	S	Std
7.5 - 1	57.5	28284	-	7.5000	7470A	PPB	05 Jun 2020 13:50:35	S	Std
10.0 - 1	57.0	38178	-	10.0000	7470A	PPB	05 Jun 2020 13:52:41	S	Std
ICV73 - 1	ICV73	15044	3.9504	-	7470A	PPB	05 Jun 2020 13:55:12	U	SMPL
ICB73 - 1	ICB73	-159	-0.0622	-	7470A	PPB	05 Jun 2020 13:57:17	U	SMPL
CCV68 - 1	CCV68	18502	4.8631	-	7470A	PPB	05 Jun 2020 13:59:23	U	SMPL
CCB68 - 1	CCB68	-149	-0.0596	-	7470A	PPB	05 Jun 2020 14:01:28	U	SMPL
CRA - 1	CRA	740	0.1751	-	7470A	PPB	05 Jun 2020 14:03:34	U	SMPL
HighStd - 1	HighStd	37836	9.9661	-	7470A	PPB	05 Jun 2020 14:05:40	U	SMPL
ChkStd - 1	ChkStd	25548	6.7229	-	7470A	PPB	05 Jun 2020 14:07:46	U	SMPL
PB129341BL - 1	PBW	-173	-0.0659	-	7470A	PPB	05 Jun 2020 14:09:54	U	SMPL
PB129341BS - 1	LCSW	14042	3.6860	-	7470A	PPB	05 Jun 2020 14:12:02	U	SMPL
L2889-01 - 1	FDW-115	21	-0.0147	-	7470A	PPB	05 Jun 2020 14:14:08	U	SMPL
L2889-01DUP - 1	FDW-115DUP	86	0.0024	-	7470A	PPB	05 Jun 2020 14:16:15	U	SMPL
L2889-01MS - 1	FDW-115MS	13297	3.4893	-	7470A	PPB	05 Jun 2020 14:18:23	U	SMPL
L2889-01MSD - 1	FDW-115MSD	15140	3.9758	-	7470A	PPB	05 Jun 2020 14:20:30	U	SMPL
CCV69 - 1	CCV69	18238	4.7935	-	7470A	PPB	05 Jun 2020 14:22:36	U	SMPL
CCB69 - 1	CCB69	-170	-0.0651	-	7470A	PPB	05 Jun 2020 14:24:42	U	SMPL
PB129341TB - 1	PB129322TB	-61	-0.0363	-	7470A	PPB	05 Jun 2020 14:26:49	U	SMPL
L2889-01LX5 - 1		-27	-0.0274	-	7470A	PPB	05 Jun 2020 14:28:55	U	SMPL
L2889-01A - 1		14927	3.9196	-	7470A	PPB	05 Jun 2020 14:31:01	U	SMPL
CCV70 - 1	CCV70	18098	4.7565	-	7470A	PPB	05 Jun 2020 14:33:07	U	SMPL
CCB70 - 1	CCB70	-175	-0.0664	-	7470A	PPB	05 Jun 2020 14:35:13	U	SMPL