

LB109462


ISM02.4

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42657



A= 0.0000e+000

B= 2.3523e-004

C= 4.2512e-003

Rho= 0.9999572

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.002	-0.002	-28	0.000	-28					
0.05	0.050					223					
0.20	0.200	0.215	0.015	894	0.0 %	894					
2.50	2.500	2.533	0.033	10751	0.0 %	10751					
5.00	5.000	4.939	-0.061	20980	0.0 %	20980					
7.50	7.500	7.477	-0.023	31769	0.0 %	31769					
10.0	10.000	10.038	0.038	42657	0.0 %	42657					

118-100

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	-28	-	0	ISM02.4	PPB	6/5/2020 15:11	Std
	0.2 S0.2	894	-	0.2	ISM02.4	PPB	6/5/2020 15:13	Std
	2.5 S2.5	10751	-	2.5	ISM02.4	PPB	6/5/2020 15:15	Std
	5 S5.0	20980	-	5	ISM02.4	PPB	6/5/2020 15:18	Std
	7.5 S7.5	31769	-	7.5	ISM02.4	PPB	6/5/2020 15:20	Std
	10 S10.0	42657	-	10	ISM02.4	PPB	6/5/2020 15:22	Std
ICV	ICV	16510	3.8878 -		ISM02.4	PPB	6/5/2020 15:25	SMPL
ICB	ICB	-104	-0.0202 -		ISM02.4	PPB	6/5/2020 15:27	SMPL
CCV071	CCV071	20816	4.9007 -		ISM02.4	PPB	6/5/2020 15:29	SMPL
CCB071	CCB071	-102	-0.0197 -		ISM02.4	PPB	6/5/2020 15:32	SMPL
PB129337BL	PBW337	-79	-0.0143 -		ISM02.4	PPB	6/5/2020 15:34	SMPL
L2845-01	MCOAE9	-130	-0.0263 -		ISM02.4	PPB	6/5/2020 15:36	SMPL
L2845-02	MCOAF6	-122	-0.0244 -		ISM02.4	PPB	6/5/2020 15:39	SMPL
L2845-03	MCOAG0	-113	-0.0223 -		ISM02.4	PPB	6/5/2020 15:41	SMPL
L2845-04	MCOAF3	104	0.0287 -		ISM02.4	PPB	6/5/2020 15:43	SMPL
L2845-05	MCOAF4	-46	-0.0066 -		ISM02.4	PPB	6/5/2020 15:45	SMPL
L2845-06	MCOAF0	-147	-0.0303 -		ISM02.4	PPB	6/5/2020 15:48	SMPL
L2845-07	MCOAF5	-90	-0.0169 -		ISM02.4	PPB	6/5/2020 15:50	SMPL
L2845-08	MCOAE5	-72	-0.0127 -		ISM02.4	PPB	6/5/2020 15:52	SMPL
L2845-09	MCOAE6	-86	-0.016 -		ISM02.4	PPB	6/5/2020 15:54	SMPL
L2845-10	MCOAE4	-127	-0.0256 -		ISM02.4	PPB	6/5/2020 15:57	SMPL
L2845-11	MCOAG1	-128	-0.0259 -		ISM02.4	PPB	6/5/2020 15:59	SMPL
L2845-12	MCOAG2	-114	-0.0226 -		ISM02.4	PPB	6/5/2020 16:01	SMPL
L2845-13	MCOAF9	-127	-0.0256 -		ISM02.4	PPB	6/5/2020 16:03	SMPL
L2845-14	MCOAF9D	-116	-0.023 -		ISM02.4	PPB	6/5/2020 16:06	SMPL
L2845-15	MCOAF9S	4067	0.9609 -		ISM02.4	PPB	6/5/2020 16:08	SMPL
L2845-16	MCOAE8	-147	-0.0303 -		ISM02.4	PPB	6/5/2020 16:10	SMPL
L2845-17	MCOAF2	-120	-0.024 -		ISM02.4	PPB	6/5/2020 16:13	SMPL
L2845-18	MCOAF7	-119	-0.0237 -		ISM02.4	PPB	6/5/2020 16:15	SMPL
L2845-19	MCOAG5	-131	-0.0266 -		ISM02.4	PPB	6/5/2020 16:17	SMPL
CCV072	CCV072	21232	4.9985 -		ISM02.4	PPB	6/5/2020 16:19	SMPL
CCB072	CCB072	-86	-0.016 -		ISM02.4	PPB	6/5/2020 16:22	SMPL
L2845-20	MCOAG6	-113	-0.0223 -		ISM02.4	PPB	6/5/2020 16:24	SMPL
L2845-21	MCOAE7	-105	-0.0204 -		ISM02.4	PPB	6/5/2020 16:26	SMPL
L2845-22	MCOAF8	-82	-0.015 -		ISM02.4	PPB	6/5/2020 16:28	SMPL
PB129350BL	PBW350	-73	-0.0129 -		ISM02.4	PPB	6/5/2020 16:31	SMPL
L2847-01	MBFTW9	-85	-0.0157 -		ISM02.4	PPB	6/5/2020 16:33	SMPL
L2847-02	MBFTX0	-71	-0.0124 -		ISM02.4	PPB	6/5/2020 16:35	SMPL
L2847-03	MBFTX1	-78	-0.0141 -		ISM02.4	PPB	6/5/2020 16:37	SMPL
L2847-04	MBFTX2	-67	-0.0115 -		ISM02.4	PPB	6/5/2020 16:40	SMPL
L2847-05	MBFTX3	-102	-0.0197 -		ISM02.4	PPB	6/5/2020 16:42	SMPL
L2847-06	MBFTX4	-109	-0.0214 -		ISM02.4	PPB	6/5/2020 16:44	SMPL
L2847-07	MBFTX5	-100	-0.0193 -		ISM02.4	PPB	6/5/2020 16:47	SMPL
L2847-08	MBFTX6	-96	-0.0183 -		ISM02.4	PPB	6/5/2020 16:49	SMPL
L2847-09	MBFTX8	-85	-0.0157 -		ISM02.4	PPB	6/5/2020 16:51	SMPL
L2847-10	MBFTX8D	-110	-0.0216 -		ISM02.4	PPB	6/5/2020 16:53	SMPL

LB109462
INSTRUMENT ID : CV1

L2847-11	MBFTX8S	3929	0.9285 -	ISM02.4	PPB	6/5/2020 16:56 SMPL
L2847-12	MBFTX9	-141	-0.0289 -	ISM02.4	PPB	6/5/2020 16:58 SMPL
L2847-13	MBFTY0	-87	-0.0162 -	ISM02.4	PPB	6/5/2020 17:00 SMPL
L2847-14	MBFTY2	-111	-0.0219 -	ISM02.4	PPB	6/5/2020 17:02 SMPL
L2847-15	MBFTY3	-105	-0.0204 -	ISM02.4	PPB	6/5/2020 17:05 SMPL
L2847-16	MBFTY4	-112	-0.0221 -	ISM02.4	PPB	6/5/2020 17:07 SMPL
CCV073	CCV073	21084	4.9637 -	ISM02.4	PPB	6/5/2020 17:09 SMPL
CCB073	CCB073	-105	-0.0204 -	ISM02.4	PPB	6/5/2020 17:11 SMPL
L2847-17	MBFTY5	-111	-0.0219 -	ISM02.4	PPB	6/5/2020 17:14 SMPL
L2847-18	MBFTY6	-108	-0.0212 -	ISM02.4	PPB	6/5/2020 17:16 SMPL
L2847-19	MBFTY7	-99	-0.019 -	ISM02.4	PPB	6/5/2020 17:18 SMPL
CCV074	CCV074	20897	4.9197 -	ISM02.4	PPB	6/5/2020 17:21 SMPL
CCB074	CCB074	-122	-0.0244 -	ISM02.4	PPB	6/5/2020 17:23 SMPL