



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109469

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/05/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	06/05/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/05/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	06/05/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.02	06/05/2020	17:36
6	L2812-05	1	Solid	20.03	20	21.6	8.22	06/05/2020	17:39
7	L2815-05	1	Solid	20.04	20	21.9	7.96	06/05/2020	17:40
8	L2886-03	1	Solid	20.02	20	22.1	10.87	06/05/2020	17:41
9	L2887-02	1	Solid	20.03	20	21.9	6.74	06/05/2020	17:42
10	L2939-05	1	Solid	20.03	20	22.3	7.65	06/05/2020	17:43
11	L2939-06	1	Solid	20.03	20	22.5	7.91	06/05/2020	17:44
12	L2939-07	1	Solid	20.02	20	22.2	7.81	06/05/2020	17:45
13	L2939-07DUP	1	Solid	20.03	20	22.1	7.82	06/05/2020	17:46
14	CCV2	1	Water	NA	NA	20.1	12.02	06/05/2020	17:47

28109469

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2890 WorkList ID : 139352 Department : Wet-Chemistry Date : 06-05-2020 09:43:31

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2020	Solid	L2812-05	Corrosivity	Cool 4 deg C	PSEG05	O41	OK-01-060520	06/05/2020	9045D
06/05/2020	Solid	L2815-05	Corrosivity	Cool 4 deg C	PSEG05	O41	OR-02-060520	06/05/2020	9045D
06/04/2020	Solid	L2886-03	Corrosivity	Cool 4 deg C	PSEG01	O21	GEN	06/04/2020	9045D
06/04/2020	Solid	L2887-02	Corrosivity	Cool 4 deg C	PSEG01	O22	CHRT-25606	06/04/2020	9045D
06/02/2020	Solid	L2939-05	Corrosivity	Cool 4 deg C	NOBI03	O63	PHC-326FB-010	06/02/2020	9045D
06/02/2020	Solid	L2939-06	Corrosivity	Cool 4 deg C	NOBI03	O63	2020-06-02-FD02	06/02/2020	9045D
06/02/2020	Solid	L2939-07	Corrosivity	Cool 4 deg C	NOBI03	O63	PHC-326FB-011	06/02/2020	9045D

Date/Time 06/05/2020 16:00
 Received by: JA
 Relinquished by: KS

Date/Time 06/05/2020 18:35
 Received by: KS
 Relinquished by: JD