



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109470

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/05/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	06/05/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/05/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	06/05/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.02	06/05/2020	17:36
6	L2923-05	1	Water	NA	NA	19.8	4.49	06/05/2020	17:50
7	L2923-06	1	Water	NA	NA	19.7	7.20	06/05/2020	17:51
8	L2923-07	1	Water	NA	NA	20.1	4.52	06/05/2020	18:00
9	L2923-10	1	Water	NA	NA	21.0	7.97	06/05/2020	18:01
10	L2925-01	1	Water	NA	NA	21.4	4.66	06/05/2020	18:02
11	L2925-02	1	Water	NA	NA	21.7	5.61	06/05/2020	18:03
12	L2925-03	1	Water	NA	NA	21.5	4.08	06/05/2020	18:05
13	L2925-04	1	Water	NA	NA	21.6	5.14	06/05/2020	18:06
14	L2925-04DUP	1	Water	NA	NA	21.7	5.15	06/05/2020	18:08
15	CCV2	1	Water	NA	NA	20.1	12.02	06/05/2020	18:10

WORKLIST(Hardcopy Internal Chain)

25109470

WorkList Name : PH L2925

WorkList ID : 139366

Department : Wet-Chemistry

Date : 06-05-2020 12:43:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2020	Water	L2923-05 ✓	pH	Cool 4 deg C	PSEG01	O42	31049	06/05/2020	9040C
06/05/2020	Water	L2923-06 ✓	pH	Cool 4 deg C	PSEG01	O42	31096	06/05/2020	9040C
06/05/2020	Water	L2923-07 ✓	pH	Cool 4 deg C	PSEG01	O42	31095	06/05/2020	9040C
06/05/2020	Water	L2923-10 ✓	pH	Cool 4 deg C	PSEG01	O42	23850	06/05/2020	9040C
06/05/2020	Water	L2925-01 ✓	pH	Cool 4 deg C	PSEG01	O31	32395	06/05/2020	9040C
06/05/2020	Water	L2925-02 ✓	pH	Cool 4 deg C	PSEG01	O31	32394	06/05/2020	9040C
06/05/2020	Water	L2925-03 ✓	pH	Cool 4 deg C	PSEG01	O31	32468	06/05/2020	9040C
06/05/2020	Water	L2925-04 ✓	pH	Cool 4 deg C	PSEG01	O31	32334	06/05/2020	9040C

Date/Time 06-05-2020 18:40
 Received by: Jh
 Relinquished by: KS

Date/Time 06-05-2020 18:35
 Received by: KS
 Relinquished by: Jh