



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109471

Slope : 98.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/05/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	7.01	06/05/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/05/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	06/05/2020	17:35
5	CCV1	1	Water	NA	NA	20.2	2.02	06/05/2020	17:36
6	L2923-01	1	Solid	20.03	20	21.7	6.34	06/05/2020	18:15
7	L2923-01DUP	1	Solid	20.04	20	21.6	6.35	06/05/2020	18:16
8	L2923-02	1	Solid	20.02	20	21.6	5.14	06/05/2020	18:19
9	CCV2	1	Water	NA	NA	20.1	12.02	06/05/2020	18:20

28109471

WORKLIST(Hardcopy Internal Chain)

WorkList Name: PH S L2923 WorkList ID: 139376 Department: Wet-Chemistry Date: 06-05-2020 17:06:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/05/2020	Solid	L2923-01 ✓	pH	Cool 4 deg C	PSEG01	O42	31008	06/05/2020	9045D
06/05/2020	Solid	L2923-02 ✓	pH	Cool 4 deg C	PSEG01	O42	31048	06/05/2020	9045D

Date/Time 06/05/2020 17:10
Received by: JS
Relinquished by: _____

Date/Time 06-05-2020 18:30
Received by: JS
Relinquished by: JD