



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109488

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/08/2020	14:25
2	CAL2	1	Water	NA	NA	20.1	7.01	06/08/2020	14:26
3	CAL3	1	Water	NA	NA	20.1	10.04	06/08/2020	14:27
4	ICV	1	Water	NA	NA	20.1	6.98	06/08/2020	14:30
5	CCV1	1	Water	NA	NA	20.2	2.01	06/08/2020	14:32
6	L2946-02	1	Water	NA	NA	20.3	7.30	06/08/2020	14:35
7	L2946-02DUP	1	Water	NA	NA	20.4	7.31	06/08/2020	14:36
8	CCV2	1	Water	NA	NA	20.1	12.02	06/08/2020	14:39

182109488

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2946 WorkList ID : 139432 Department : Wet-Chemistry Date : 06-08-2020 14:19:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2020	Water	L2946-02	-X pH	Cool 4 deg C	PSEG01	O42	PURGE	06/08/2020	9040C

Date/Time 06/08/2020 14:20
Received by: Jo
Relinquished by: Jo

Date/Time 06/08/2020 15:35
Received by: Jo
Relinquished by: Jo