



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 6/9/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 06/08/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 06/09/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB109496

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L2943-01	1-1	1	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-02	1-2	2	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-03	2-3	3	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-04	2-4	4	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-05	3-5	5	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-06	3-6	6	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-07	4-7	7	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-08	4-8	8	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-09	5-9	9	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2943-10	5-10	10	1.00	2.00	2.00	100.0	CAULKING SAMPLE,100% SOLIDS
L2946-05	CUTTINGS	11	1.14	9.72	8.51	85.9	
L2946-07	POLY	12	1.00	2.00	2.00	100.0	POLY SAMPLE,100% SOLIDS
L2948-01	SYNERTECH	13	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L2948-03	SONOTUBE	14	1.00	2.00	2.00	100.0	STONE SAMPLE,100% SOLIDS
L2950-01	72-11939	15	1.13	9.84	7.33	71.2	
L2951-01	239988	16	1.16	9.96	9.52	95.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$