


LB109501

7470A

INSTRUMENT ID: CV 2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	132	0.000	132				
0.2	0.200	0.195	-0.005	788	0.0 %	788				
2.5	2.500	2.454	-0.046	9811	0.0 %	9811				
5.0	5.000	4.919	-0.081	19656	0.0 %	19656				
7.5	7.500	7.730	0.230	30887	0.0 %	30887				
10.0	10.000	9.872	-0.128	39439	0.0 %	39439				

LB109501 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 09 Jun 2020 08:43:10					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	132	-	0.0000	7470A	PPB	09 Jun 2020 09:33:24	S	Std
0.2 - 1	50.2	788	-	0.2000	7470A	PPB	09 Jun 2020 09:35:32	S	Std
2.5 - 1	52.5	9811	-	2.5000	7470A	PPB	09 Jun 2020 09:38:11	S	Std
5.0 - 1	55	19656	-	5.0000	7470A	PPB	09 Jun 2020 09:40:19	S	Std
7.5 - 1	57.5	30887	-	7.5000	7470A	PPB	09 Jun 2020 09:46:41	S	Std
10.0 - 1	510	39439	-	10.0000	7470A	PPB	09 Jun 2020 09:48:48	S	Std
ICV75 - 1	ICV75	15342	3.8386	-	7470A	PPB	09 Jun 2020 09:51:40	U	SMPL
ICB75 - 1	ICB75	45	0.0088	-	7470A	PPB	09 Jun 2020 09:53:46	U	SMPL
CCV80 - 1	CCV80	19837	4.9640	-	7470A	PPB	09 Jun 2020 09:57:26	U	SMPL
CCB80 - 1	CCB80	57	0.0118	-	7470A	PPB	09 Jun 2020 09:59:32	U	SMPL
CRA - 1	CRA	793	0.1961	-	7470A	PPB	09 Jun 2020 10:01:38	U	SMPL
HighStd - 1	HighStd	40143	10.0478	-	7470A	PPB	09 Jun 2020 10:03:45	U	SMPL
ChkStd - 1	ChkStd	28136	7.0417	-	7470A	PPB	09 Jun 2020 10:05:51	U	SMPL
PB129398BL - 1	PBW	-161	-0.0427	-	7470A	PPB	09 Jun 2020 10:10:32	U	SMPL
PB129398BS - 1	LCSW	15226	3.8096	-	7470A	PPB	09 Jun 2020 10:12:38	U	SMPL
L2812-05 - 1	OK-01-060520	480	0.1178	-	7470A	PPB	09 Jun 2020 10:14:45	U	SMPL
L2812-05DUP - 1	OK-01-060520DUP	465	0.1140	-	7470A	PPB	09 Jun 2020 10:16:53	U	SMPL
L2812-05MS - 1	OK-01-060520MS	15554	3.8917	-	7470A	PPB	09 Jun 2020 10:19:01	U	SMPL
L2812-05MSD - 1	OK-01-060520MSD	14860	3.7179	-	7470A	PPB	09 Jun 2020 10:21:09	U	SMPL
L2815-05 - 1	OR-02-060520	140	0.0326	-	7470A	PPB	09 Jun 2020 10:23:15	U	SMPL
CCV81 - 1	CCV81	19890	4.9772	-	7470A	PPB	09 Jun 2020 10:25:22	U	SMPL
CCB81 - 1	CCB81	51	0.0103	-	7470A	PPB	09 Jun 2020 10:27:28	U	SMPL
L2886-03 - 1	GEN	190	0.0451	-	7470A	PPB	09 Jun 2020 10:29:34	U	SMPL
L2887-02 - 1	CHRT-25606	573	0.1410	-	7470A	PPB	09 Jun 2020 10:31:41	U	SMPL
L2890-02 - 1	CHRT-26160	103	0.0234	-	7470A	PPB	09 Jun 2020 10:33:47	U	SMPL
L2923-03 - 1	COMP-2	305	0.0739	-	7470A	PPB	09 Jun 2020 10:35:54	U	SMPL
L2939-05 - 1	PHC-326FB-010	271	0.0654	-	7470A	PPB	09 Jun 2020 10:38:01	U	SMPL
L2939-06 - 1	2020-06-02-FD02	462	0.1132	-	7470A	PPB	09 Jun 2020 10:40:08	U	SMPL
L2939-07 - 1	PHC-326FB-011	403	0.0985	-	7470A	PPB	09 Jun 2020 10:42:16	U	SMPL
PB129399BL - 1	PBW	-29	-0.0097	-	7470A	PPB	09 Jun 2020 10:44:24	U	SMPL
PB129399BS - 1	LCSW	15030	3.7605	-	7470A	PPB	09 Jun 2020 10:46:32	U	SMPL
L2923-08 - 1	COMP-1	240	0.0577	-	7470A	PPB	09 Jun 2020 10:48:39	U	SMPL
CCV82 - 1	CCV82	19624	4.9106	-	7470A	PPB	09 Jun 2020 10:50:45	U	SMPL
CCB82 - 1	CCB82	58	0.0121	-	7470A	PPB	09 Jun 2020 10:52:52	U	SMPL
L2923-08DUP - 1	COMP-1DUP	134	0.0311	-	7470A	PPB	09 Jun 2020 10:55:00	U	SMPL
L2923-08MS - 1	COMP-1MS	15533	3.8864	-	7470A	PPB	09 Jun 2020 10:57:06	U	SMPL
L2923-08MSD - 1	COMP-1MSD	15112	3.7810	-	7470A	PPB	09 Jun 2020 10:59:13	U	SMPL
L2923-10 - 1	23850	112	0.0256	-	7470A	PPB	09 Jun 2020 11:01:21	U	SMPL
PB129400BL - 1	PBW	-94	-0.0260	-	7470A	PPB	09 Jun 2020 11:03:29	U	SMPL
PB129400BS - 1	LCSW	15023	3.7587	-	7470A	PPB	09 Jun 2020 11:05:36	U	SMPL
L2926-01 - 1	TANK-1-20200604	13	0.0008	-	7470A	PPB	09 Jun 2020 11:07:44	U	SMPL
L2926-01DUP - 1	TANK-1-20200604DUP	135	0.0314	-	7470A	PPB	09 Jun 2020 11:09:53	U	SMPL
L2926-01MS - 1	TANK-1-20200604MS	14566	3.6443	-	7470A	PPB	09 Jun 2020 11:12:01	U	SMPL
L2926-01MSD - 1	TANK-1-20200604MSD	13478	3.3719	-	7470A	PPB	09 Jun 2020 11:14:08	U	SMPL
CCV83 - 1	CCV83	20007	5.0065	-	7470A	PPB	09 Jun 2020 11:16:15	U	SMPL
CCB83 - 1	CCB83	35	0.0063	-	7470A	PPB	09 Jun 2020 11:18:22	U	SMPL
L2946-02 - 1	PURGE	24	0.0036	-	7470A	PPB	09 Jun 2020 11:20:29	U	SMPL
PB129407BL - 1	PBW	174	0.0411	-	7470A	PPB	09 Jun 2020 11:22:35	U	SMPL
PB129407BS - 1	LCSW	14819	3.7077	-	7470A	PPB	09 Jun 2020 11:24:42	U	SMPL
L2941-02 - 1	RBF-TS-001	289	0.0699	-	7470A	PPB	09 Jun 2020 11:26:49	U	SMPL
L2941-02DUP - 1	RBF-TS-001DUP	114	0.0261	-	7470A	PPB	09 Jun 2020 11:30:42	U	SMPL
L2941-02MS - 1	RBF-TS-001MS	14137	3.5369	-	7470A	PPB	09 Jun 2020 11:32:50	U	SMPL
L2941-02MSD - 1	RBF-TS-001MSD	13982	3.4981	-	7470A	PPB	09 Jun 2020 11:34:57	U	SMPL
PB129398TB - 1	PB12939TB	17	0.0018	-	7470A	PPB	09 Jun 2020 11:39:44	U	SMPL
PB129399TB - 1	PB129391TB	18	0.0021	-	7470A	PPB	09 Jun 2020 11:43:45	U	SMPL
PB129407TB - 1	PB129393TB	9	-0.0002	-	7470A	PPB	09 Jun 2020 11:45:52	U	SMPL
CCV84 - 1	CCV84	19855	4.9685	-	7470A	PPB	09 Jun 2020 11:47:58	U	SMPL
CCB84 - 1	CCB84	81	0.0179	-	7470A	PPB	09 Jun 2020 11:50:05	U	SMPL
L2812-05LX5 - 1		217	0.0519	-	7470A	PPB	09 Jun 2020 11:52:12	U	SMPL
L2812-05A - 1		15536	3.8872	-	7470A	PPB	09 Jun 2020 11:54:19	U	SMPL
L2923-08LX5 - 1		108	0.0246	-	7470A	PPB	09 Jun 2020 11:56:26	U	SMPL
L2923-08A - 1		15181	3.7983	-	7470A	PPB	09 Jun 2020 11:58:33	U	SMPL
L2926-01LX5 - 1		65	0.0139	-	7470A	PPB	09 Jun 2020 12:00:40	U	SMPL
L2926-01A - 1		17510	4.3814	-	7470A	PPB	09 Jun 2020 12:02:47	U	SMPL
L2941-02LX5 - 1		110	0.0251	-	7470A	PPB	09 Jun 2020 12:04:55	U	SMPL
L2941-02A - 1		17595	4.4027	-	7470A	PPB	09 Jun 2020 12:07:03	U	SMPL
CCV85 - 1	CCV85	19878	4.9742	-	7470A	PPB	09 Jun 2020 12:09:11	U	SMPL
CCB85 - 1	CCB85	76	0.0166	-	7470A	PPB	09 Jun 2020 12:11:18	U	SMPL