



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109502

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/08/2020	17:30
2	CAL2	1	Water	NA	NA	20.1	6.99	06/08/2020	17:31
3	CAL3	1	Water	NA	NA	20.1	10.03	06/08/2020	17:32
4	ICV	1	Water	NA	NA	20.1	7.00	06/08/2020	17:33
5	CCV1	1	Water	NA	NA	20.1	2.01	06/08/2020	17:35
6	L2946-06	1	Solid	20.03	20	21.5	8.20	06/08/2020	17:37
7	L2950-02	1	Solid	20.02	20	21.8	7.86	06/08/2020	17:40
8	L2951-02	1	Solid	20.03	20	21.4	7.72	06/08/2020	17:42
9	L2951-02DUP	1	Solid	20.04	20	21.5	7.73	06/08/2020	17:44
10	CCV2	1	Water	NA	NA	20.2	12.02	06/08/2020	17:45

WORKLIST(Hardcopy Internal Chain)

28 109502

WorkList Name : CORROSIONITY L2946 S(WorkList ID : 139435 Department : Wet-Chemistry Date : 06-08-2020 14:20:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2020	Solid	L2946-06	Corrosivity	Cool 4 deg C	PSEG01	O42	CUTTINGS	06/08/2020	9045D
06/08/2020	Solid	L2950-02	Corrosivity	Cool 4 deg C	PSEG01	O42	72-11939	06/08/2020	9045D
06/08/2020	Solid	L2951-02	Corrosivity	Cool 4 deg C	PSEG01	O43	239988	06/08/2020	9045D

Date/Time 06-08-2020 16:40
 Received by: JP
 Relinquished by: KS

Date/Time 06-08-2020 18:00
 Received by: KS
 Relinquished by: JP