



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109508

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/09/2020	11:15
2	CAL2	1	Water	NA	NA	20.1	7.01	06/09/2020	11:16
3	CAL3	1	Water	NA	NA	20.1	10.03	06/09/2020	11:17
4	ICV	1	Water	NA	NA	20.2	6.98	06/09/2020	11:20
5	CCV1	1	Water	NA	NA	20.1	2.01	06/09/2020	11:22
6	L2949-01	1	Water	NA	NA	22.1	6.04	06/09/2020	11:24
7	L2949-02	1	Water	NA	NA	17.3	6.47	06/09/2020	11:25
8	L2949-03	1	Water	NA	NA	17.6	6.46	06/09/2020	11:26
9	L2949-04	1	Water	NA	NA	21.2	6.59	06/09/2020	11:29
10	L2949-04DUP	1	Water	NA	NA	21.1	6.60	06/09/2020	11:30
11	CCV2	1	Water	NA	NA	20.1	12.02	06/09/2020	11:32

WORKLIST(Hardcopy Internal Chain)

LB 109508

WorkList Name : PH L2949

WorkList ID : 139465

Department : Wet-Chemistry

Date : 06-09-2020 11:04:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2949-01	pH	Cool 4 deg C	LOCK01		MW-1	06/04/2020	SM 4500-H B
06/04/2020	Water	L2949-02	pH	Cool 4 deg C	LOCK01		MW-3	06/04/2020	SM 4500-H B
06/04/2020	Water	L2949-03	pH	Cool 4 deg C	LOCK01		MW-4	06/04/2020	SM 4500-H B
06/04/2020	Water	L2949-04	pH	Cool 4 deg C	LOCK01		MW-2	06/04/2020	SM 4500-H B

Date/Time 06-09-2020 11:10

Received by: Jn

Relinquished by: PM

Date/Time 06-09-2020 12:30

Received by: AM

Relinquished by: Jn