



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB109512

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP84723
Turbidity Calibration std, 20NTU	WP84716
Turbidity Calibration std, 80NTU	WP84714
Turbidity Calibration std, 40NTU	WP84715
Turbidity Calibration std, 1NTU	WP84719
Turbidity Calibration std, 5NTU	WP84718
Turbidity Calibration - CCV std, 10 NTU	WP84717
10 NTU Standard 500 ml	W2489

Intercept: 0.1481

Slope: 0.966

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.289	06/09/2020	12:15
2	CAL2	1	1	0.985	06/09/2020	12:17
3	CAL3	5	1	4.860	06/09/2020	12:19
4	CAL4	10	1	10.200	06/09/2020	12:21
5	CAL5	20	1	19.400	06/09/2020	12:24
6	CAL6	40	1	38.400	06/09/2020	12:26
7	CAL7	80	1	77.600	06/09/2020	12:28



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB109512

ANALYST: Alexander
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.400	10.613	06/09/2020	12:32
2	ICB		1	0.277	0.133	06/09/2020	12:35
3	CCV1	10	1	10.100	10.302	06/09/2020	12:38
4	CCB1		1	0.290	0.147	06/09/2020	12:40
5	LB109512BL		1	0.282	0.139	06/09/2020	12:42
6	L2949-01		1	1.220	1.110	06/09/2020	12:44
7	L2949-01DUP		1	1.300	1.192	06/09/2020	12:46
8	L2949-02		1	2.420	2.352	06/09/2020	12:48
9	L2949-03		1	2.010	1.927	06/09/2020	12:50
10	L2949-04		1	63.200	65.271	06/09/2020	12:52
11	CCV2	10	1	10.200	10.406	06/09/2020	12:55
12	CCB2		1	0.293	0.150	06/09/2020	12:58

LB109512 WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity-6092020 WorkList ID : 139472 Department : Wet-Chemistry Date : 06-09-2020 11:55:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/06/2020	Water	L2949-01	H Turbidity	Cool 4 deg C	LOCK01	MW-1		06/04/2020	SM2130 B
06/06/2020	Water	L2949-02	H Turbidity	Cool 4 deg C	LOCK01	MW-3		06/04/2020	SM2130 B
06/06/2020	Water	L2949-03	H Turbidity	Cool 4 deg C	LOCK01	MW-4		06/04/2020	SM2130 B
06/06/2020	Water	L2949-04	H Turbidity	Cool 4 deg C	LOCK01	MW-2		06/04/2020	SM2130 B

Date/Time 6/09/2020 12:00
 Received by: JP AM
 Relinquished by: JP

Date/Time 6/09/2020 13:05
 Received by: JP AM
 Relinquished by: JP