


LB109517

7471B

INSTRUMENT ID: CVL



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.051	0.051	107	0.000	107				
0.2	0.200	0.210	0.010	1043	0.0 %	1043				
2.5	2.500	2.474	-0.026	14434	0.0 %	14434				
5.0	5.000	4.984	-0.016	29280	0.0 %	29280				
7.5	7.500	7.378	-0.122	43437	0.0 %	43437				
10.0	10.000	10.103	0.103	59554	0.0 %	59554				

LB109517 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Jun 2020 14:35:58

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.0000	7471B	PPB	09 Jun 2020 14:38:19	S	Std
0.2 - 1	502	1043	-	0.2000	7471B	PPB	09 Jun 2020 14:40:36	S	Std
2.5 - 1	525	14434	-	2.5000	7471B	PPB	09 Jun 2020 14:42:54	S	Std
5.0 - 1	55	29280	-	5.0000	7471B	PPB	09 Jun 2020 14:45:11	S	Std
7.5 - 1	575	43437	-	7.5000	7471B	PPB	09 Jun 2020 14:47:28	S	Std
10.0 - 1	595	59554	-	10.0000	7471B	PPB	09 Jun 2020 14:49:45	S	Std
ICV76 - 1	ICV76	22579	3.8511	-	7471B	PPB	09 Jun 2020 14:52:55	U	SMPL
ICB76 - 1	ICB76	79	0.0467	-	7471B	PPB	09 Jun 2020 14:55:10	U	SMPL
CCV86 - 1	CCV86	29252	4.9794	-	7471B	PPB	09 Jun 2020 14:57:26	U	SMPL
CCB86 - 1	CCB86	-34	0.0276	-	7471B	PPB	09 Jun 2020 14:59:42	U	SMPL
CRA - 1	CRA	1011	0.2043	-	7471B	PPB	09 Jun 2020 15:01:58	U	SMPL
HighStd - 1	HighStd	60071	10.1904	-	7471B	PPB	09 Jun 2020 15:04:14	U	SMPL
ChkStd - 1	ChkStd	40130	6.8187	-	7471B	PPB	09 Jun 2020 15:06:30	U	SMPL
PB129436BL - 1	PBS	-249	-0.0087	-	7471B	PPB	09 Jun 2020 15:08:47	U	SMPL
PB129436BS - 1	LCSS	19463	3.3242	-	7471B	PPB	09 Jun 2020 15:11:04	U	SMPL
L2946-05 - 1	CUTTINGS	74086	12.5601	-	7471B	PPB	09 Jun 2020 15:18:31	U	SMPL
L2946-05DUP - 1	CUTTINGSDUP	71562	12.1333	-	7471B	PPB	09 Jun 2020 15:20:48	U	SMPL
L2946-05MS - 1	CUTTINGSMS	82539	13.9894	-	7471B	PPB	09 Jun 2020 15:23:07	U	SMPL
L2946-05MSD - 1	CUTTINGSMSD	78612	13.3254	-	7471B	PPB	09 Jun 2020 15:28:16	U	SMPL
L2950-01 - 1	72-11939	5157	0.9053	-	7471B	PPB	09 Jun 2020 15:30:31	U	SMPL
CCV87 - 1	CCV87	29482	5.0183	-	7471B	PPB	09 Jun 2020 15:32:49	U	SMPL
CCB87 - 1	CCB87	70	0.0452	-	7471B	PPB	09 Jun 2020 15:35:04	U	SMPL
L2951-01 - 1	239988	4061	0.7200	-	7471B	PPB	09 Jun 2020 15:37:20	U	SMPL
L2946-05LX5 - 1		16771	2.8691	-	7471B	PPB	09 Jun 2020 15:41:26	U	SMPL
L2946-05A - 1		83828	14.2073	-	7471B	PPB	09 Jun 2020 15:43:41	U	SMPL
L2946-05DLX2 - 1		34399	5.8497	-	7471B	PPB	09 Jun 2020 15:45:57	U	SMPL
L2946-05DUPDLX2 - 1		33923	5.7692	-	7471B	PPB	09 Jun 2020 15:48:16	U	SMPL
L2946-05MSDLX2 - 1		36997	6.2889	-	7471B	PPB	09 Jun 2020 15:50:33	U	SMPL
L2946-05MSDDLX2 - 1		38900	6.6107	-	7471B	PPB	09 Jun 2020 15:55:45	U	SMPL
L2946-05LDLX10 - 1		7969	1.3808	-	7471B	PPB	09 Jun 2020 15:58:02	U	SMPL
L2946-05ADLX2 - 1		40757	6.9247	-	7471B	PPB	09 Jun 2020 16:00:20	U	SMPL
CCV88 - 1	CCV88	29521	5.0249	-	7471B	PPB	09 Jun 2020 16:02:36	U	SMPL
CCB88 - 1	CCB88	7	0.0345	-	7471B	PPB	09 Jun 2020 16:04:53	U	SMPL