


LB109533

INSTRUMENT ID: CV1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.025	0.025	141	0.000	141				
0.2	0.200	0.191	-0.009	822	0.0 %	822				
2.5	2.500	2.482	-0.018	10268	0.0 %	10268				
5.0	5.000	4.905	-0.095	20258	0.0 %	20258				
7.5	7.500	7.706	0.206	31810	0.0 %	31810				
10.0	10.000	9.891	-0.109	40818	0.0 %	40818				

LB109533 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 10 Jun 2020 09:41:40					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.0000	7470A	PPB	10 Jun 2020 09:45:41	S	Std
0.2 - 1	50.2	822	-	0.2000	7470A	PPB	10 Jun 2020 09:47:49	S	Std
2.5 - 1	52.5	10268	-	2.5000	7470A	PPB	10 Jun 2020 09:49:57	S	Std
5.0 - 1	55	20258	-	5.0000	7470A	PPB	10 Jun 2020 09:52:05	S	Std
7.5 - 1	57.5	31810	-	7.5000	7470A	PPB	10 Jun 2020 09:54:13	S	Std
10.0 - 1	59	40818	-	10.0000	7470A	PPB	10 Jun 2020 09:56:19	S	Std
ICV77 - 1	ICV77	16637	4.0264	-	7470A	PPB	10 Jun 2020 09:59:02	U	SMPL
ICB77 - 1	ICB77	-156	-0.0466	-	7470A	PPB	10 Jun 2020 10:01:08	U	SMPL
CCV89 - 1	CCV89	20281	4.9102	-	7470A	PPB	10 Jun 2020 10:03:14	U	SMPL
CCB89 - 1	CCB89	-162	-0.0480	-	7470A	PPB	10 Jun 2020 10:05:21	U	SMPL
CRA - 1	CRA	828	0.1921	-	7470A	PPB	10 Jun 2020 10:07:27	U	SMPL
HighStd - 1	HighStd	40679	9.8575	-	7470A	PPB	10 Jun 2020 10:09:34	U	SMPL
ChkStd - 1	ChkStd	27512	6.6640	-	7470A	PPB	10 Jun 2020 10:11:40	U	SMPL
PB129442BL - 1	PBW	-185	-0.0536	-	7470A	PPB	10 Jun 2020 10:13:48	U	SMPL
PB129442BS - 1	LCSW	13867	3.3546	-	7470A	PPB	10 Jun 2020 10:15:56	U	SMPL
L2946-06 - 1	CUTTINGS	-130	-0.0403	-	7470A	PPB	10 Jun 2020 10:18:03	U	SMPL
L2946-06DUP - 1	CUTTINGSDUP	-66	-0.0247	-	7470A	PPB	10 Jun 2020 10:20:11	U	SMPL
L2946-06MS - 1	CUTTINGSMS	14677	3.5510	-	7470A	PPB	10 Jun 2020 10:22:19	U	SMPL
L2946-06MSD - 1	CUTTINGSMSD	15165	3.6694	-	7470A	PPB	10 Jun 2020 10:24:28	U	SMPL
L2950-02 - 1	72-11939	-130	-0.0403	-	7470A	PPB	10 Jun 2020 10:26:34	U	SMPL
CCV90 - 1	CCV90	20290	4.9124	-	7470A	PPB	10 Jun 2020 10:28:40	U	SMPL
CCB90 - 1	CCB90	-172	-0.0504	-	7470A	PPB	10 Jun 2020 10:30:46	U	SMPL
L2951-02 - 1	239988	-86	-0.0296	-	7470A	PPB	10 Jun 2020 10:32:53	U	SMPL
PB129442TB - 1	PB129409TB	-74	-0.0267	-	7470A	PPB	10 Jun 2020 10:34:59	U	SMPL
L2946-06LX5 - 1		101	0.0158	-	7470A	PPB	10 Jun 2020 10:37:06	U	SMPL
L2946-06A - 1		16800	4.0659	-	7470A	PPB	10 Jun 2020 10:39:13	U	SMPL
CCV91 - 1	CCV91	20344	4.9255	-	7470A	PPB	10 Jun 2020 10:41:20	U	SMPL
CCB91 - 1	CCB91	-163	-0.0483	-	7470A	PPB	10 Jun 2020 10:43:27	U	SMPL