



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109537

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/10/2020	10:40
2	CAL2	1	Water	NA	NA	20.1	7.00	06/10/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	06/10/2020	10:42
4	ICV	1	Water	NA	NA	20.2	7.00	06/10/2020	10:44
5	CCV1	1	Water	NA	NA	20.1	2.01	06/10/2020	10:45
6	L2967-01	1	Water	NA	NA	21.8	4.89	06/10/2020	10:46
7	L2967-01DUP	1	Water	NA	NA	21.9	4.90	06/10/2020	10:47
8	L2969-01	1	Water	NA	NA	21.3	4.46	06/10/2020	10:49
9	CCV2	1	Water	NA	NA	20.1	12.02	06/10/2020	10:50

LB 108537

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2967 WorkList ID : 139497 Department : Wet-Chemistry Date : 06-10-2020 10:33:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/04/2020	Water	L2967-01 1	pH	Cool 4 deg C	STER15	O52	GILROY-OMEGA-CELL	06/04/2020	SM 4500-H B
06/08/2020	Water	L2969-01, 2	pH	Cool 4 deg C	STER13	O52	CELL-B-POOL-WATER	06/08/2020	SM 4500-H B

Date/Time 06.10.2020 10:35
Received by: Jo
Relinquished by: AM

Date/Time 06.10.2020 11:45
Received by: AM
Relinquished by: Jo