



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109538

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/10/2020	10:40
2	CAL2	1	Water	NA	NA	20.1	7.00	06/10/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	06/10/2020	10:42
4	ICV	1	Water	NA	NA	20.2	7.00	06/10/2020	10:44
5	CCV1	1	Water	NA	NA	20.2	2.02	06/10/2020	12:30
6	L2971-01	1	Water	NA	NA	22.0	10.73	06/10/2020	12:32
7	L2971-01DUP	1	Water	NA	NA	22.1	10.74	06/10/2020	12:33
8	L2971-03	1	Water	NA	NA	21.8	9.60	06/10/2020	12:35
9	CCV2	1	Water	NA	NA	20.1	12.02	06/10/2020	12:36

18109538

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2971 WorkList ID : 139500 Department : Wet-Chemistry Date : 06-10-2020 12:23:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/10/2020	Water	L2971-01	pH	Cool 4 deg C	PSEG01	O52	30698	06/10/2020	9040C
06/10/2020	Water	L2971-03	pH	Cool 4 deg C	PSEG01	O52	40755	06/10/2020	9040C

Date/Time 06.10.2020 12:25
 Received by: Ja
 Relinquished by: Ja

Date/Time 06.10.2020 13:30
 Received by: Ja
 Relinquished by: Ja