



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB109553

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/10/2020	17:35
2	CAL2	1	Water	NA	NA	20.2	7.00	06/10/2020	17:36
3	CAL3	1	Water	NA	NA	20.2	10.03	06/10/2020	17:37
4	ICV	1	Water	NA	NA	20.2	7.00	06/10/2020	17:39
5	CCV1	1	Water	NA	NA	20.2	2.01	06/10/2020	17:40
6	L2810-08	1	Solid	20.03	20	22.7	8.67	06/10/2020	17:42
7	L2817-05	1	Solid	20.02	20	24.0	7.91	06/10/2020	17:43
8	L2819-03	1	Solid	20.03	20	21.7	8.66	06/10/2020	17:45
9	L2970-03	1	Solid	20.02	20	21.7	8.24	06/10/2020	17:46
10	L2970-03DUP	1	Solid	20.03	20	21.8	8.25	06/10/2020	17:49
11	CCV2	1	Water	NA	NA	20.2	12.02	06/10/2020	17:50

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY L2972

WorkList ID : 139509

Department : Wet-Chemistry

Date : 06-10-2020 13:44:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/09/2020	Solid	L2810-08	Corrosivity	Cool 4 deg C	PSEG05	O51	CL-02-060920	06/09/2020	9045D
06/10/2020	Solid	L2817-05	Corrosivity	Cool 4 deg C	PSEG05	O52	SU-03-061020	06/10/2020	9045D
06/10/2020	Solid	L2819-03	Corrosivity	Cool 4 deg C	PSEG05	O53	BU-02-061020	06/10/2020	9045D
06/10/2020	Solid	L2970-03	Corrosivity	Cool 4 deg C	PSEG01	O52	600472986	06/10/2020	9045D

Date/Time 06.10.2020 17:15
 Received by: Ja
 Relinquished by: Dr

Date/Time 06.10.2020 18:30
 Received by: Sh
 Relinquished by: JD