



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109554

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	06/11/2020	10:15
2	CAL2	1	Water	NA	NA	20.1	6.99	06/11/2020	10:16
3	CAL3	1	Water	NA	NA	20.4	10.03	06/11/2020	10:17
4	ICV	1	Water	NA	NA	20.1	7.00	06/11/2020	10:19
5	CCV1	1	Water	NA	NA	20.2	2.01	06/11/2020	10:20
6	L2974-05	1	Water	NA	NA	15.2	7.88	06/11/2020	10:21
7	L2974-05DUP	1	Water	NA	NA	15.3	7.89	06/11/2020	10:22
8	CCV2	1	Water	NA	NA	20.2	12.02	06/11/2020	10:25

WORKLIST(Hardcopy Internal Chain)

WB 109554

WorkList Name : PHWL2974 WorkList ID : 139543 Department : Wet-Chemistry Date : 06-11-2020 10:04:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/10/2020	Water	L2974-05	pH	Cool 4 deg C	PSEG01	O53	30986	06/10/2020	9040C

Date/Time 06.11.2020 10:10
 Received by: Jp
 Relinquished by: KS

Date/Time 06.11.2020 11:25
 Received by: KS
 Relinquished by: Jp