



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB109561

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	06/11/2020	10:05
2	L2946-03		1	8	100.00	>212.0	>212.0	06/11/2020	10:35
3	L2946-03DUP		1	8	100.00	>212.0	>212.0	06/11/2020	11:10
4	L2971-01		1	9	100.00	>212.0	>212.0	06/11/2020	11:40
5	L2971-03		1	9	100.00	>212.0	>212.0	06/11/2020	12:05
6	L2974-05		1	8	100.00	>212.0	>212.0	06/11/2020	12:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB109561

WorkList Name : fp61120 WorkList ID : 139539 Department : Wet-Chemistry Date : 06-11-2020 09:23:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2020	Water	L2946-03	C Flash Point	Cool 4 deg C	PSEG01	O42	PURGE	06/08/2020	1010A
06/16/2020	Water	L2971-01	C Flash Point	Cool 4 deg C	PSEG01	O52	30698	06/10/2020	1010A
06/16/2020	Water	L2971-03	C Flash Point	Cool 4 deg C	PSEG01	O52	40755	06/10/2020	1010A
06/23/2020	Water	L2974-05	C Flash Point	Cool 4 deg C	PSEG01	O53	30986	06/10/2020	1010A

Date/Time 06/11/2020 09:30
 Received by: KS
 Relinquished by: JS

Date/Time 06/11/2020 12:40
 Received by: JS
 Relinquished by: KS