



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB109566

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2626
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	06/11/2020	15:25
2	CAL2	1	Water	NA	NA	20.2	7.01	06/11/2020	15:26
3	CAL3	1	Water	NA	NA	20.2	10.03	06/11/2020	15:27
4	ICV	1	Water	NA	NA	20.2	7.00	06/11/2020	15:30
5	CCV1	1	Water	NA	NA	20.2	2.01	06/11/2020	15:32
6	L2986-01	1	Water	NA	NA	18.6	6.56	06/11/2020	15:33
7	L2986-01DUP	1	Water	NA	NA	18.7	6.57	06/11/2020	15:35
8	L2986-02	1	Water	NA	NA	19.9	6.61	06/11/2020	15:36
9	L2986-03	1	Water	NA	NA	18.6	6.62	06/11/2020	15:37
10	L2986-04	1	Water	NA	NA	18.9	6.58	06/11/2020	15:40
11	L2986-05	1	Water	NA	NA	17.7	6.65	06/11/2020	15:42
12	L2986-06	1	Water	NA	NA	23.6	6.07	06/11/2020	15:43
13	L2986-07	1	Water	NA	NA	22.7	5.77	06/11/2020	15:45
14	CCV2	1	Water	NA	NA	20.1	12.02	06/11/2020	15:47

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L2986 WorkList ID : 139559 Department : Wet-Chemistry Date : 06-11-2020 15:15:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2020	Water	L2986-01 ✓	pH	Cool 4 deg C	PSEG01	P21	1416	06/11/2020	9040C
06/11/2020	Water	L2986-02 ✓	pH	Cool 4 deg C	PSEG01	P21	1417	06/11/2020	9040C
06/11/2020	Water	L2986-03 ✓	pH	Cool 4 deg C	PSEG01	P21	1418	06/11/2020	9040C
06/11/2020	Water	L2986-04 ✓	pH	Cool 4 deg C	PSEG01	P21	1419	06/11/2020	9040C
06/11/2020	Water	L2986-05 ✓	pH	Cool 4 deg C	PSEG01	P21	1420	06/11/2020	9040C
06/11/2020	Water	L2986-06 ✓	pH	Cool 4 deg C	PSEG01	P21	1393	06/11/2020	9040C
06/11/2020	Water	L2986-07 ✓	pH	Cool 4 deg C	PSEG01	P21	1394	06/11/2020	9040C

Date/Time 06/11/2020 15:20
 Received by: JS
 Relinquished by: JS

Date/Time 06/11/2020 17:00
 Received by: JS
 Relinquished by: JS